#### Spectroscopy and Such (Working Title)

By Blaise Jonathan Thompson

A dissertation submitted in partial fulfillment of the requirements for the degree of

Doctor of Philosophy (Chemistry)

at the UNIVERSITY OF WISCONSIN - MADISON  $$20\ensuremath{\mathrm{xx}}$$ 

Date of final oral examination: xx/xx/xxxx

This dissertation is approved by the following members of the Final Oral Committee: John C. Wright, Professor, Analytical Chemistry

Х

Х

Х

Χ

# **Contents**

| List of Figures List of Tables |        |   |          |  |
|--------------------------------|--------|---|----------|--|
|                                |        |   |          |  |
| ΑI                             | bstrac | ct  |          |  |
| 1                              | Intr   | oduction  | 1        |  |
|                                | 1.1    | Coherent Multidimensional Spectroscopy            | 1        |  |
|                                | 1.2    | The CMDS Instrument                               | 1        |  |
|                                | 1.3    | Scientific Software                               | 2        |  |
| •                              | D      |   | •        |  |
| l                              |        | ckground  | 3        |  |
| 2                              | -      | ctroscopy   | 5        |  |
|                                | 2.1    | Light   | 5        |  |
|                                | 2.2    | Light-Matter Interaction                          | 5        |  |
|                                |        | 2.2.1 Representations                             | 5        |  |
|                                | 2.3    | Linear Spectroscopy                               | 7        |  |
|                                |        | 2.3.1 Reflectivity                                | 7        |  |
|                                | 2.4    | Coherent Multidimensional Spectroscopy            | 7        |  |
|                                |        | 2.4.1 Three Wave                                  | 7<br>7   |  |
|                                |        | 2.4.3 Five Wave                                   | 8        |  |
|                                |        | 2.4.4 Six Wave                                    | 8        |  |
|                                | 2.5    | Strategies for CMDS                               | 8        |  |
|                                |        | 2.5.1 Homodyne vs. Heterodyne Detection           | 8        |  |
|                                |        | 2.5.2 Frequency vs. Time Domain                   | 8        |  |
|                                |        | 2.5.3 Triply Electronically Enhanced Spectroscopy | 9        |  |
|                                |        | 2.5.4 Transient Absorbance Spectroscopy           | 9        |  |
|                                |        | 2.5.5 Cross Polarized TrEE                        | 13<br>13 |  |
|                                | 0.6    | ·   |          |  |
|                                | 2.6    | Instrumental Response Function                    | 13       |  |

|    |      | 2.6.1 Time Domain       1         2.6.2 Frequency Domain       1         2.6.3 Time-Bandwidth Product       1 | 4 |
|----|------|---|---|
| 3  | Mat  | erials  |   |
| 4  |      | entangling material and instrument response   |   |
| -  | 4.1  | Introduction  |   |
|    | 4.2  | Theory  |   |
|    |      | 4.2.1 Inhomogeneity   |   |
|    | 4.3  | Methods   | 9 |
|    |      | 4.3.1 Characteristics of Driven and Impulsive Response  |   |
|    |      | 4.3.2 Convolution Technique for Inhomogeneous Broadening  |   |
|    | 4.4  | Results   |   |
|    |      | 4.4.1 Evolution of single coherence   |   |
|    |      | 4.4.3 Temporal pathway discrimination   |   |
|    |      | 4.4.4 Multidimensional line shape dependence on pulse delay time 4  |   |
|    |      | 4.4.5 Inhomogeneous broadening  | 4 |
|    | 4.5  | Discussion  | 9 |
|    |      | 4.5.1 An intuitive picture of pulse effects   |   |
|    |      | 4.5.2 Conditional validity of the driven limit  |   |
|    |      | 4.5.3 Extracting true material correlation  |   |
|    | 4.6  | Conclusion  | ' |
| II | Ins  | strumental Development 69   | 9 |
| 5  | Soft | eware 7   | 1 |
|    | 5.1  | Overview  | 2 |
|    | 5.2  | WrightTools   | 2 |
|    | 5.3  | PyCMDS  | 2 |
|    |      | 5.3.1 Overview  |   |
|    |      | 5.3.2 Ideal Axis Positions  |   |
|    | - 4  | 5.3.3 Exponential   |   |
|    | 5.4  | WrightSim   |   |
| 6  | Inst | rumental Development 7  | 9 |
|    | 6.1  | Hardware  | - |
|    | 6.2  | Signal Acquisition  | - |
|    | 6.3  | Artifacts and Noise   |   |
|    | 0.5  | 6.3.1 Scatter   |   |

|        |                  | 6.3.2 Normalization of dual-chopped self-heterodyned signal | 87         |
|--------|------------------|---|------------|
|        | 6.4              | Light Generation  | 88         |
|        |                  | 6.4.1 Automated OPA Tuning                                  | 88         |
|        | 6.5              | Optomechanics   |            |
|        |                  | 6.5.1 Automated Neutral Density Wheels                      | 88         |
|        |                  |   |            |
|        |                  | unliantiona   | 89         |
| <br> - |                  | pplications   | 91         |
| 7      | PbS              |   |            |
| 8      | MX:              |   | 93         |
|        | 8.1              | Introduction  |            |
|        | 8.2              | Methods   |            |
|        | 8.3              | Results and discussion                                      |            |
|        | 8.4              | Conclusions   |            |
| 9      | PEC              |   | 121        |
|        | 9.1              | Introduction  |            |
|        | 9.2              | Background  | 121        |
|        | 9.3              | Methods   | 122        |
|        | 9.4              | Transmittance and reflectance                               | 123        |
|        | 9.5              | Three-pulse echo spectroscopy                               |            |
|        |                  | 9.5.1 Assignment of zero delay                              |            |
|        | 9.6              | Frequency-domain transient grating spectroscopy             | 140        |
| 10     | Pyri             | te  | 141        |
| 11     | BiV              | 04  | 143        |
|        |                  |   |            |
|        |                  |   |            |
| I۷     | ′ A <sub>l</sub> | ppendix 1   | <b>L45</b> |
| Α      | Pub              | lic :   | 147        |
|        | A.1              | Chemical systems  | 147        |
|        |                  | A.1.1 Concentration   | 148        |
|        |                  | A.1.2 Timescale   |            |
|        |                  | A.1.3 Lengthscale   |            |
|        | A.2              | Analytical chemistry  |            |
|        | A.3              | Spectroscopy  |            |
|        |                  | A.3.1 Nonlinear spectroscopy                                |            |
|        | A.4              | Instrumentation   |            |
|        |                  | A.4.1 LASER   |            |
|        |                  |   |            |

| В | Prod | cedures 1                      | 151 |
|---|------|--------------------------------|-----|
|   | B.1  | Aligining TOPAS-C              | 151 |
|   | B.2  | Aligning Spitfire PRO          | 151 |
|   | B.3  | Air Handling                   | 151 |
|   | B.4  | Six Month Maintenance          | 151 |
|   | B.5  | Tuning MicroHR Monochromator   | 151 |
| C | Hard | dware 1                        | 153 |
|   | C.1  | Adjustable periscopes          |     |
|   | C.2  | Automated transmissive filters | 155 |
|   | C.3  | Electronics                    | 156 |
| D | Erra | ta 1                           | 157 |
|   | D.1  | Czech 2015                     | 157 |
| Е | Colo | pphon 1                        | 159 |

# **List of Figures**

| 2.1  | CAPTION TODO   | 10 |
|------|--|----|
| 4.1  | Sixteen triply-resonant Liouville pathways.  | 23 |
| 4.2  | Overview of the MR-CMDS simulation.  | 26 |
| 4.3  | Convolution overview   | 33 |
| 4.4  | Relative importance of FID and driven response for a single quantum coherence        | 37 |
| 4.5  | Pulsed excitation of a single quantum coherence and its dependance on pulse detuning | 39 |
| 4.6  | 2D frequency response of a single Liouville pathway at different delay values        | 42 |
| 4.7  | 2D delay response for different relative dephasing rates                             | 47 |
| 4.8  | Evolution of the 2D frequency response   | 50 |
| 4.9  | Wigners  | 53 |
| 4.10 | 2D delay response with inhomogeneity   | 55 |
| 4.11 | Spectral evolution of an inhomogenious system  | 58 |
| 4.12 | Conditional validity of the driven limit   | 62 |
| 4.13 | Metrics of correlation   | 64 |
| 5.1  | TODO   | 77 |
| 6.1  | Simulated interference paterns in old delay parameterization                         | 81 |
| 6.2  | Simulated interference paterns in current delay parameterization                     | 83 |
| 6.3  | Comparison of single, dual chopping  | 86 |

| 8.1  | CMDS tutorial  |
|------|--|
| 8.2  | Schemiatic of the synthetic setup used for Mo thin film sulfidation reactions 98 |
| 8.3  | Mask and epi vs transmissive   |
| 8.4  | OPA outputs at each color explored   |
| 8.5  | Spectral delay correction  |
| 8.6  | MoS <sub>2</sub> post processing   |
| 8.7  | Few-layer $MoS_2$ thin film characterization                                     |
| 8.8  | $MoS_2$ absorbance   |
| 8.9  | $MoS_2$ frequency-frequency slices   |
| 8.10 | $MoS_2 \omega_1$ Wigner progression  |
| 8.11 | $MoS_2 \omega_2$ Wigner progression  |
| 8.12 | Pathway V, VI liouville pathways   |
| 8.13 | $MoS_2$ transients   |
| 8.14 | $MoS_2$ frequency-frequency slices near pulse overlap                            |
| 8.15 | Pathways I, III Liouville pathways   |
| 9.1  | PEDOT:PSS transmission and reflectance spectra                                   |
| 9.2  | PEDOT:PSS 3PE phase matching mask  |
| 9.3  | PEDOT:PSS 3PE raw data   |
| 9.4  | PEDOT:PSS 3PE delay space  |
| 9.5  | PEDOT:PSS 3PE processed data   |
| 9.6  | PEDOT:PSS 3PE traces   |
| 9.7  | PEDOT:PSS 3PE traces   |
| 9.8  | PEDOT:PSS 3PE traces   |
| 9.9  | PEDOT:PSS 3PE traces   |
| C.1  | CAPTION TODO   |

# **List of Tables**

| 4.1 | Conditions for peak intensity at different pulse delays for pathway I $\gamma$ | 44  |
|-----|--|-----|
| 6.1 | Shot-types in phase shifted parallel modulation                                | 84  |
| 9.1 |  | 138 |

# Acknowledgments

## **Abstract**

The explanatory stories that people find compelling are simple; are concrete rather than abstract; assign a larger role to talent, stupidity and intentions than to luck; and focus on a few striking events that happened rather than on the countless events that failed to happen.

The ultimate test of an explanation is whether it would have made the event predictable in advance.

Paradoxically, it is easier to construct a coherent story when you know little, when there are fewer pieces to fit into the puzzle. Our comforting conviction that the world makes sense rests on a secure foundation: our almost unlimited ability to ignore our ignorance.

- Daniel Kahneman [1]

## Chapter 1

## Introduction

### 1.1 Coherent Multidimensional Spectroscopy

CMDS, coherent multidimensional spectroscopy

#### 1.2 The CMDS Instrument

From an instrumental perspective, MR-CMDS is a problem of calibration and coordination. Within the Wright Group, each of our two main instruments are composed of roughly ten actively moving component hardwares. Many of these components are purchased directly from vendors such as SpectraPhysics, National Instruments, Horiba, Thorlabs, and Newport. Others are created or heavily modified by graduate students. The Wright Group has always maintained custom acquisition software packages which control the complex, many-stepped dance that these components must perform to acquire MR-CMDS spectra.

#### 1.3 Scientific Software

When I joined the Wright Group, I saw that acquisition software was a real barrier to experimental progress and flexibility. Graduate students had ideas for instrumental enhancements that were infeasible because of the challenge of incorporating the new components into the existing software ecosystem. At the same time, students were spending much of their time in lab repeatedly calibrating optical parametric amplifiers by hand, a process that sometimes took days. I chose to spend a significant portion of my graduate career focusing on solving these problems through software development. At first, I focused on improving the existing LabVIEW code. Eventually, I developed a vision for a deeply modular acquisition software that could not be practically created with LabVIEW. Using Python and Qt, I created a brand new acquisition software PyCMDS: built from the ground up to fundamentally solve historical challenges in the Group. PyCMDS offers a modular hardware model that can "re-configure" itself to flexibly control a variety of component hardware configurations. This has enabled graduate students to add and remove hardware whenever necessary, without worrying about a heavy additional programming burden. PyCMDS is now used to drive both MR-CMDS instruments in the Group, allowing for easy sharing of component hardware and lessening the total amount of software that the Group needs to maintain. Besides being more flexible, PyCMDS solves a number of other problems. It offers fully automated strategies for calibrating component hardwares, making calibration less arduous and more reproducible. It offers more fine-grained control of data acquisition and timing, enabling more complex algorithms to quickly acquire artifact-free results. In conjunction with other algorithmic and hardware improvements that I have made, PyCMDS has decreased acquisition times by up to two orders of magnitude. A companion software, WrightTools (which I also created), solves some of the processing and representation challenges of multidimensional data.

## Part I

**Background** 

## Chapter 2

## **Spectroscopy**

In this chapter I lay out the foundations of spectroscopy.

## 2.1 Light

## 2.2 Light-Matter Interaction

Spectroscopic experiments all derive from the interaction of light and matter. Many material properties can be deduced by measuring the nature of this interaction.

Nonlinear spectroscopy relies upon higher-order terms in the light-matter interaction. In a generic system, each term is roughly ten times smaller than the last.

#### 2.2.1 Representations

Many strategies have been introduced for diagrammatically representing the interaction of multiple electric fields in an experiment.

6

#### Circle Diagrams

#### **Double-sided Feynman Diagrams**

#### **WMEL Diagrams**

So-called wave mixing energy level (WMEL) diagrams are the most familiar way of representing spectroscopy for Wright group members. WMEL diagrams were first proposed by Lee and Albrecht in an appendix to their seminal work *A Unified View of Raman, Resonance Raman, and Fluorescence Spectroscopy* [2]. WMEL diagrams are drawn using the following rules.

- 1. The energy ladder is represented with horizontal lines solid for real states and dashed for virtual states.
- 2. Individual electric field interactions are represented as vertical arrows. The arrows span the distance between the initial and final state in the energy ladder.
- 3. The time ordering of the interactions is represented by the ordering of arrows, from left to right.
- 4. Ket-side interactions are represented with solid arrows.
- 5. Bra-side interactions are represented with dashed arrows.
- 6. Output is represented as a solid wavy line.

#### **Mukamel Diagrams**

## 2.3 Linear Spectroscopy

#### 2.3.1 Reflectivity

This derivation adapted from *Optical Processes in Semiconductors* by Jacques I. Pankove [3]. For normal incidence, the reflection coefficient is

$$R = \frac{(n-1)^2 + k^2}{(n+1)^2 + k^2} \tag{2.1}$$

Further derivation adapted from [4]. To extend reflectivity to a differential measurement

## 2.4 Coherent Multidimensional Spectroscopy

multiresonant coherent multidimensional spectroscopy

#### 2.4.1 Three Wave

#### 2.4.2 Four Wave

Fluorescence

Raman

#### 2.4.3 Five Wave

#### 2.4.4 Six Wave

multiple population-period transient spectroscopy (MUPPETS)

#### 2.5 Strategies for CMDS

#### 2.5.1 Homodyne vs. Heterodyne Detection

Two kinds of spectroscopies: 1) heterodyne 2) homodyne. Heterodyne techniques may be self heterodyne or explicitly heterodyned with a local oscillator.

In all heterodyne spectroscopies, signal goes as N. In all homodyne spectroscopies, signal goes as  $N^2$ . This literally means that homodyne signals go as the square of heterodyne signals, which is what we mean when we say that homodyne signals are intensity level and heterodyne signals are amplitude level.

Transient absorption, TA

#### 2.5.2 Frequency vs. Time Domain

Time domain techniques become more and more difficult when large frequency bandwidths are needed. With very short, broad pulses:

- Non-resonant signal becomes brighter relative to resonant signal
- Pulse distortions become important.

This epi-CARS paper might have some useful discussion of non-resonant vs resonant for shorter and shorter pulses [5].

An excellent discussion of pulse distortion phenomena in broadband time-domain experiments was published by Spencer et al. [6].

Another idea in defense of frequency domain is for the case of power studies. Since time-domain pulses in-fact possess all colors in them they cannot be trusted as much at perturbative fluence. See that paper that Natalia presented...

#### 2.5.3 Triply Electronically Enhanced Spectroscopy

Triply Electronically Enhanced (TrEE) spectroscopy has become the workhorse homodyne-detected 4WM experiment in the Wright Group.

#### 2.5.4 Transient Absorbance Spectroscopy

Transient absorption (TA)

#### Quantitative TA

Transient absorbance (TA) spectroscopy is a self-heterodyned technique. Through chopping you can measure nonlinearities quantitatively much easier than with homodyne detected (or explicitly heterodyned) experiments.




Figure 2.1: CAPTION TODO

Figure 2.1 diagrams the TA measurement for a generic sample. Here I show measurement of both the reflected and transmitted probe beam ... not important in opaque (pyrite) or non-reflective (quantum dot) samples ...

Typically one attempts to calculate the change in absorbance  $\Delta A \dots$ 

$$\Delta A = A_{\rm on} - A_{\rm off} \tag{2.2}$$

$$= -\log_{10}\left(\frac{I_{T} + I_{R} + I_{\Delta T} + I_{\Delta R}}{I_{0}}\right) + \log\left(\frac{I_{T} + I_{R}}{I_{0}}\right)$$
(2.3)

$$= -(\log_{10}(I_{\mathsf{T}} + I_{\mathsf{R}} + I_{\Delta\mathsf{T}} + I_{\Delta\mathsf{R}}) - \log_{10}(I_{\mathsf{0}})) + (\log_{10}(I_{\mathsf{T}} + I_{\mathsf{R}}) - \log_{10}(I_{\mathsf{0}}))$$
(2.4)

$$= -(\log_{10}(I_{\mathsf{T}} + I_{\mathsf{R}} + I_{\Delta\mathsf{T}} + I_{\Delta\mathsf{R}}) - \log_{10}(I_{\mathsf{T}} + I_{\mathsf{R}})) \tag{2.5}$$

$$= -\log_{10}\left(\frac{I_{\mathsf{T}} + I_{\mathsf{R}} + I_{\Delta\mathsf{T}} + I_{\Delta\mathsf{R}}}{I_{\mathsf{T}} + I_{\mathsf{R}}}\right) \tag{2.6}$$

Equation 2.6 simplifies beautifully if reflectivity is negligible . . .

Now I define a variable for each experimental measurable:

 $V_{
m T}$  voltage recorded from transmitted beam, without pump  $V_{
m R}$  voltage recorded from reflected beam, without pump  $V_{
m \Delta T}$  change in voltage recorded from transmitted beam due to pump  $V_{
m \Delta R}$  change in voltage recorded from reflected beam due to pump

We will need to calibrate using a sample with a known transmisivity and reflectivity constant:

$$V_{\text{T, ref}}$$
 voltage recorded from transmitted beam, without pump  $V_{\text{R, ref}}$  voltage recorded from reflected beam, without pump  $\mathcal{T}_{\text{ref}}$  transmissivity  $\mathcal{R}_{\text{ref}}$  reflectivity

Define two new proportionality constants...

$$C_{\rm T} \equiv \frac{\mathcal{T}}{V_{\rm T}}$$
 (2.7)  
 $C_{\rm R} \equiv \frac{\mathcal{R}}{V_{\rm R}}$ 

$$C_{\mathsf{R}} \equiv \frac{\mathcal{R}}{V_{\mathsf{R}}} \tag{2.8}$$

These are explicitly calibrated (as a function of probe color) prior to the experiment using the calibration sample.

Given the eight experimental measurables ( $V_T$ ,  $V_R$ ,  $V_{\Delta T}$ ,  $V_{\Delta R}$ ,  $V_{T, ref}$ ,  $V_{R, ref}$ ,  $\mathcal{T}_{ref}$ ,  $\mathcal{R}_{ref}$ ) I can express all of the intensities in Equation 2.6 in terms of  $I_0$ .

$$C_{\mathsf{T}} = \frac{\mathcal{T}_{\mathsf{ref}}}{V_{\mathsf{T},\mathsf{ref}}} \tag{2.9}$$

$$C_{T} = \frac{\mathcal{T}_{ref}}{V_{T, ref}}$$

$$C_{R} = \frac{\mathcal{R}_{ref}}{V_{R, ref}}$$
(2.9)

$$I_{\mathsf{T}} = I_0 C_{\mathsf{T}} V_{\mathsf{T}} \tag{2.11}$$

$$I_{\mathsf{R}} = I_0 C_{\mathsf{R}} V_{\mathsf{R}} \tag{2.12}$$

$$I_{\Delta \mathsf{T}} = I_0 C_{\mathsf{T}} V_{\Delta \mathsf{T}} \tag{2.13}$$

$$I_{\Delta R} = I_0 C_R V_{\Delta R} \tag{2.14}$$

Wonderfully, the  $I_0$  cancels when plugged back in to Equation 2.6, leaving a final expression for  $\Delta A$ that only depends on my eight measurables.

$$\Delta A = -\log_{10} \left( \frac{C_{T}(V_{T} + V_{\Delta T}) + C_{R}(V_{R} + V_{\Delta R})}{C_{T}V_{T} + C_{R}V_{R}} \right)$$
(2.15)

#### 2.5.5 **Cross Polarized TrEE**

#### 2.5.6 Pump-TrEE-Probe

Pump TrEE probe (PTP).

#### 2.6 **Instrumental Response Function**

The instrumental response function (IRF) is a classic concept in analytical science. Defining IRF becomes complex with instruments as complex as these, but it is still useful to attempt.

It is particularly useful to define bandwidth.

#### 2.6.1 **Time Domain**

I will use four wave mixing to extract the time-domain pulse-width. I use a driven signal e.g. near infrared carbon tetrachloride response. I'll homodyne-detect the output. In my experiment I'm moving pulse 1 against pulses 2 and 3 (which are coincident).

The driven polarization, P, goes as the product of my input pulse intensities:

$$P(T) = I_1(t - T) \times I_2(t) \times I_3(t)$$
 (2.16)

In our experiment we are convolving  $l_1$  with  $l_2 \times l_3$ . Each pulse has an *intensity-level* width,  $\sigma_1$ ,  $\sigma_2$ , and  $\sigma_3$ .  $I_2 \times I_3$  is itself a Gaussian, and

$$\sigma_{I_2I_3} = \dots (2.17)$$

$$\sigma_{I_2I_3} = \dots$$

$$= \sqrt{\frac{\sigma_2^2 \sigma_3^2}{\sigma_2^2 + \sigma_3^2}}.$$
(2.17)

The width of the polarization (across T) is therefore

$$\sigma_P = \sqrt{\sigma_1^2 + \sigma_{l_2 l_3}^2} (2.19)$$

$$= \dots (2.20)$$

$$= \dots$$

$$= \sqrt{\frac{\sigma_1^2 + \sigma_2^2 \sigma_3^2}{\sigma_1^2 + \sigma_2^2}}.$$
(2.20)

I assume that all of the pulses have the same width.  $I_1$ ,  $I_2$ , and  $I_3$  are identical Gaussian functions with FWHM  $\sigma$ . In this case, Equation 2.21 simplifies to

$$\sigma_P = \sqrt{\frac{\sigma^2 + \sigma^2 \sigma^2}{\sigma^2 + \sigma^2}} \tag{2.22}$$

$$= \dots (2.23)$$

$$= \sigma\sqrt{\frac{3}{2}} \tag{2.24}$$

Finally, since we measure  $\sigma_P$  and wish to extract  $\sigma$ :

$$\sigma = \sigma_P \sqrt{\frac{2}{3}} \tag{2.25}$$

Again, all of these widths are on the intensity level.

#### 2.6.2 **Frequency Domain**

We can directly measure  $\sigma$  (the width on the intensity-level) in the frequency domain using a spectrometer. A tune test contains this information.

#### 2.6.3 Time-Bandwidth Product

For a Gaussian, approximately 0.441

## **Chapter 3**

## **Materials**

"Kroemer's Lemma of Proven Ignorance": If, in discussing a semiconductor problem, you cannot draw an Energy Band Diagram, this shows that you don't know what you are talking about, If you can draw one, but don't, then your audience won't know what you are talking about.

## Chapter 4

# Disentangling material and instrument response

Ultrafast spectroscopy is often collected in the mixed frequency/time domain, where pulse durations are similar to system dephasing times. In these experiments, expectations derived from the familiar driven and impulsive limits are not valid. This work simulates the mixed-domain Four Wave Mixing response of a model system to develop expectations for this more complex field-matter interaction. We explore frequency and delay axes. We show that these line shapes are exquisitely sensitive to excitation pulse widths and delays. Near pulse overlap, the excitation pulses induce correlations which resemble signatures of dynamic inhomogeneity. We describe these line shapes using an intuitive picture that connects to familiar field-matter expressions. We develop strategies for distinguishing pulse-induced correlations from true system inhomogeneity. These simulations provide a foundation for interpretation of ultrafast experiments in the mixed domain.

#### 4.1 Introduction

Ultrafast spectroscopy is based on using nonlinear interactions, created by multiple ultrashort ( $10^{-9} - 10^{-15}$ s) pulses, to resolve spectral information on timescales as short as the pulses themselves. [Rentzepis1970,

Mukamel2000 The ultrafast specta can be collected in the time domain or the frequency domain. [Park1998]

Time-domain methods scan the pulse delays to resolve the free induction decay (FID). [Gallagher1998] The Fourier Transform of the FID gives the ultrafast spectrum. Ideally, these experiments are performed in the impulsive limit where FID dominates the measurement. FID occurs at the frequency of the transition that has been excited by a well-defined, time-ordered sequence of pulses. Time-domain methods are compromised when the dynamics occur on faster time scales than the ultrafast excitation pulses. As the pulses temporally overlap, FID from other pulse time-orderings and emission driven by the excitation pulses both become important. These factors are responsible for the complex "coherent artifacts" that are often ignored in pump-probe and related methods. [Lebedev2007, Vardeny1981, Joffre1988, Pollard1992] Dynamics faster than the pulse envelopes are best measured using line shapes in frequency domain methods.

Frequency-domain methods scan pulse frequencies to resolve the ultrafast spectrum directly. [Druet1979, Oudar1980] Ideally, these experiments are performed in the driven limit where the steady state dominates the measurement. In the driven limit, all time-orderings of the pulse interactions are equally important and FID decay is negligible. The output signal is driven at the excitation pulse frequencies during the excitation pulse width. Frequency-domain methods are compromised when the spectral line shape is narrower than the frequency bandwidth of the excitation pulses. Dynamics that are slower than the pulse envelopes can be measured in the time domain by resolving the phase oscillations of the output signal during the entire FID decay.

There is also the hybrid mixed-time/frequency-domain approach, where pulse delays and pulse frequencies are both scanned to measure the system response. This approach is uniquely suited for experiments where the dephasing time is comparable to the pulse durations, so that neither frequency-domain nor time-domain approaches excel on their own.[Oudar1980, Wright1997a, Wright1991] In this regime, both FID and driven processes are important.[Pakoulev2006] Their relative importance depends on pulse frequencies and delays. Extracting the correct spectrum from the measurement then requires a more complex analysis that explicitly treats the excitation pulses and the different time-orderings.[Pakoulev2007, Kohler2014, Gelin2009a] Despite these complications, mixed-domain methods have a practical advantage: the dual frequency- and delay-scanning capabilities allow these

methods to address a wide variety of dephasing rates.

The relative importance of FID and driven processes and the changing importance of different coherence pathways are important factors for understanding spectral features in all ultrafast methods.

These methods include partially-coherent methods involving intermediate populations such as pump-probe, [Hamm2000] transient grating, [Salcedo1978, Fourkas1992, Fourkas1992a] transient absorption/reflection, [Au Bakker2002] photon echo, [DeBoeij1996, Patterson1984, Tokmakoff1995] two dimensional-infrared spectroscopy (2D-IR), [Hamm1999, Asplund2000, Zanni2001] 2D-electronic spectroscopy (2D-ES), [Hybl2001a, Brixner2004] and three pulse photon echo peak shift (3PEPS) [Emde1998, DeBoeij1996, DeBoeij1995, Cho1992, Passino1997] spectroscopies. These methods also include fully-coherent methods involving only coherences such as Stimulated Raman Spectroscopy (SRS), [Yoon2005, McCamant2005] Doubly Vibrationally Enhanced (DOVE), [Zhao1999, Zhao1999a, Zhao2000, Meyer2003, Donaldson2007, Donaldson2008, Fournier2008] Triply Resonant Sum Frequency (TRSF), [Boyle2013a, Boyle2013, Boyle2014] Sum Frequency Generation (SFG) [Lagutchev2007], Coherent Anti-Stokes Raman Spectroscopy (CARS) [Carlson1990b, Carlson1990a, Carlson1991] and other coherent Raman methods [Steehler1985].

This paper focuses on understanding the nature of the spectral changes that occur in Coherent Multidimensional Spectroscopy (CMDS) as experiments transition between the two limits of frequency- and time-domain methods. CMDS is a family of spectroscopies that use multiple delay and/or frequency axes to extract homogeneous and inhomogeneous broadening, as well as detailed information about spectral diffusion and chemical changes.[Kwac2003, Wright2016] For time-domain CMDS (2D-IR, 2D-ES), the complications that occur when the impulsive approximation does not strictly hold has only recently been addressed.[Erlik2017, Smallwood2016]

Frequency-domain CMDS methods, referred to herein as multi-resonant CMDS (MR-CMDS), have similar capabilities for measuring homogeneous and inhomogeneous broadening. Although these experiments are typically described in the driven limit, [Gallagher1998, Fourkas1992, Fourkas1992a] many of the experiments involve pulse widths that are comparable to the widths of the system. [Meyer2003, Donaldson2007, Pakoulev2009, Zhao1999, Czech2015, Kohler2014] MR-CMDS then becomes a mixed-domain experiment whereby resonances are characterized with marginal resolution in both frequency and time. For example, DOVE spectroscopy involves three different pathways [Wright2003]

whose relative importance depends on the relative importance of FID and driven responses.[**Donaldson2010**] In the driven limit, the DOVE line shape depends on the difference between the first two pulse frequencies so the line shape has a diagonal character that mimics the effects of inhomogeneous broadening. In the FID limit where the coherence frequencies are defined instead by the transition, the diagonal character is lost. Understanding these effects is crucial for interpreting experiments, yet these effects have not been characterized for MR-CMDS.

This work considers the third-order MR-CMDS response of a 3-level model system using three ultrafast excitation beams with the commonly used four-wave mixing (FWM) phase-matching condition,  $\vec{k}_{\text{out}} =$  $ec{k}_1-ec{k}_2+ec{k}_{2'}.$  Here, the subscripts represent the excitation pulse frequencies,  $\omega_1$  and  $\omega_2=\omega_{2'}.$  These experimental conditions were recently used to explore line shapes of excitonic systems, Kohler 2014, Czech2015] and have been developed on vibrational states as well.[Meyer2004] Although MR-CMDS forms the context of this model, the treatment is quite general because the phase matching condition can describe any of the spectroscopies mentioned above with the exception of SFG and TRSF, for which the model can be easily extended. We numerically simulate the MR-CMDS response with pulse durations at, above, and below the system coherence time. To highlight the role of pulse effects, we build an interpretation of the full MR-CMDS response by first showing how finite pulses affect the evolution of a coherence, and then how finite pulses affect an isolated third-order pathway. When considering the full MR-CMDS response, we show that spectral features change dramatically as a function of delay, even for a homogeneous system with elementary dynamics. Importantly, the line shape can exhibit correlations that mimic inhomogeneity, and the temporal evolution of this line shape can mimic spectral diffusion. We identify key signatures that can help differentiate true inhomogeneity and spectral diffusion from these measurement artifacts.

# 4.2 Theory




Figure 4.1: The sixteen triply-resonant Liouville pathways for the third-order response of the system used here. Time flows from left to right. Each excitation is labeled by the pulse stimulating the transition; excitatons with  $\omega_1$  are yellow, excitations with  $\omega_2 = \omega_{2'}$  are purple, and the final emission is gray.

We consider a simple three-level system (states n=0,1,2) that highlights the multidimensional line shape changes resulting from choices of the relative dephasing and detuning of the system and the temporal and spectral widths of the excitation pulses. For simplicity, we will ignore population relaxation effects:  $\Gamma_{11} = \Gamma_{00} = 0$ .

The electric field pulses,  $\{E_I\}$ , are given by:

$$E_{l}(t;\omega_{l},\tau_{l},\vec{k}_{l}\cdot z) = \frac{1}{2} \left[ c_{l}(t-\tau_{l})e^{i\vec{k}_{l}\cdot z}e^{-i\omega_{l}(t-\tau_{l})} + c.c. \right], \tag{4.1}$$

where  $\omega_l$  is the field carrier frequency,  $\vec{k}_l$  is the wavevector,  $\tau_l$  is the pulse delay, and  $c_l$  is a slowly varying envelope. In this work, we assume normalized (real-valued) Gaussian envelopes:

$$c_l(t) = \frac{1}{\Delta_t} \sqrt{\frac{2 \ln 2}{2\pi}} \exp\left(-\ln 2 \left[\frac{t}{\Delta_t}\right]^2\right), \tag{4.2}$$

where  $\Delta_t$  is the temporal FWHM of the envelope intensity. We neglect non-linear phase effects such as chirp so the FWHM of the frequency bandwidth is transform limited:  $\Delta_\omega \Delta_t = 4 \ln 2 \approx 2.77$ , where  $\Delta_\omega$  is the spectral FWHM (intensity scale).

The Liouville-von Neumann Equation propagates the density matrix,  $\rho$ :

$$\frac{d\rho}{dt} = -\frac{i}{\hbar} \left[ \mathbf{H_0} + \mu \cdot \sum_{l=1,2,2'} E_l(t), \rho \right] + \mathbf{\Gamma} \rho. \tag{4.3}$$

Here  $\mathbf{H_0}$  is the time-independent Hamiltonian,  $\mu$  is the dipole superoperator, and  $\Gamma$  contains the pure dephasing rate of the system. We perform the standard perturbative expansion of Equation 4.3 to third order in the electric field interaction[mukamel1995principles, Yee1978, Oudar1980, Armstrong1962, Schweigert2008] and restrict ourselves only to the terms that have the correct spatial wave vector  $\vec{k}_{\text{out}} = \vec{k}_1 - \vec{k}_2 + \vec{k}_{2'}$ . This approximation narrows the scope to sets of three interactions, one from each field, that result in the correct spatial dependence. The set of three interactions have 3! = 6 unique time-ordered sequences, and each time-ordering produces either two or three unique system-field interactions for our system, for a total of sixteen unique system-field interaction sequences, or Liouville pathways, to consider. Fig. 4.1 shows these sixteen pathways as Wave Mixing Energy Level (WMEL)

#### diagrams[Lee1985].

We first focus on a single interaction in these sequences, where an excitation pulse, x, forms  $\rho_{ij}$  from  $\rho_{ik}$  or  $\rho_{kj}$ . For brevity, we use  $\hbar=1$  and abbreviate the initial and final density matrix elements as  $\rho_i$  and  $\rho_f$ , respectively. Using the natural frequency rotating frame,  $\tilde{\rho}_{ij}=\rho_{ij}e^{-i\omega_{ij}t}$ , the formation of  $\rho_f$  using pulse x is written as

$$\frac{d\tilde{\rho}_f}{dt} = -\Gamma_f \tilde{\rho}_f 
+ \frac{i}{2} \lambda_f \mu_f c_x (t - \tau_x) e^{i\kappa_f (\vec{k}_x \cdot z + \omega_x \tau_x)} e^{i\kappa_f \Omega_{fx} t} \tilde{\rho}_i(t),$$
(4.4)

where  $\Omega_{fx}=\kappa_f^{-1}\omega_f-\omega_x(=|\omega_f|-\omega_x)$  is the detuning,  $\omega_f$  is the transition frequency of the  $i^{th}$  transition,  $\mu_f$  is the transition dipole, and  $\Gamma_f$  is the dephasing/relaxation rate for  $\rho_f$ . The  $\lambda_f$  and  $\kappa_f$  parameters describe the phases of the interaction:  $\lambda_f=+1$  for ket-side transitions and -1 for bra-side transitions, and  $\kappa_f$  depends on whether  $\rho_f$  is formed via absorption ( $\kappa_f=\lambda_f$ ) or emission ( $\kappa_f=-\lambda_f$ ). In the following equations we neglect spatial dependence (z=0).

Equation 4.4 forms the basis for our simulations. It provides a general expression for arbitrary values of the dephasing rate and excitation pulse bandwidth. The integral solution is

$$\tilde{\rho}_{f}(t) = \frac{i}{2} \lambda_{f} \mu_{f} e^{i\kappa_{f}\omega_{x}\tau_{x}} e^{i\kappa_{f}\Omega_{fx}t}$$

$$\times \int_{-\infty}^{\infty} c_{x}(t - u - \tau_{x}) \tilde{\rho}_{i}(t - u) \Theta(u)$$

$$\times e^{-(\Gamma_{f} + i\kappa_{f}\Omega_{fx})u} du,$$
(4.5)

where  $\Theta$  is the Heaviside step function. Equation 4.5 becomes the steady state limit expression when  $\Delta_t |\Gamma_f + i\kappa_f \Omega_{fx}| \gg 1$ , and the impulsive limit expression results when  $\Delta_t |\Gamma_f + i\kappa_f \Omega_{fx}| \ll 1$ . Both limits are important for understanding the multidimensional line shape changes discussed in this paper. The steady state and impulsive limits of Equation 4.5 are discussed in Appendix 4.3.1.

 $<sup>^{1}\</sup>kappa_{f}$  also has a direct relationship to the phase matching relationship: for transitions with  $E_{2}$ ,  $\kappa_{f}=1$ , and for  $E_{1}$  or  $E_{2'}$ ,  $\kappa_{f}=-1$ .




Figure 4.2: Overview of the MR-CMDS simulation. (a) The temporal profile of a coherence under pulsed excitation depends on how quickly the coherence dephases. In all subsequent panes, the relative dephasing rate is kept constant at  $\Gamma_{10}\Delta_t=1$ . (b) Simulated evolution of the density matrix elements of a third-order Liouville pathway  $V\gamma$  under fully resonant excitation. Pulses can be labeled both by their time of arrival (x,y,z) and by the lab lasers used to stimulate the transitions (2,2',1). The final coherence (teal) creates the output electric field. (c) The frequency profile of the output electric field is filtered by a monochromator gating function,  $M(\omega)$ , and the passed components (shaded) are measured. (d-f) Signal is viewed against two laser parameters, either as 2D delay (d), mixed delay-frequency (e), or 2D frequency plots (f). The six time-orderings are labeled in (d) to help introduce our delay convention.

Fig. 4.2 gives an overview of the simulations done in this work. Fig. 4.2a shows an excitation pulse (gray-shaded) and examples of a coherent transient for three different dephasing rates. The color bindings to dephasing rates introduced in Fig. 4.2a will be used consistently throughout this work. Our simulations use systems with dephasing rates quantified relative to the pulse duration:  $\Gamma_{10}\Delta_t=0.5,1,$  or 2. The temporal axes are normalized to the pulse duration,  $\Delta_t$ . The  $\Gamma_{10}\Delta_t=2$  transient is mostly driven by the excitation pulse while  $\Gamma_{10}\Delta_t=0.5$  has a substantial free induction decay (FID) component at late times. Fig. 4.2b shows a pulse sequence (pulses are shaded orange and pink) and the resulting system evolution of pathway  $V\gamma$  (00  $\stackrel{2}{\rightarrow}$  01  $\stackrel{2'}{\rightarrow}$  11  $\stackrel{1}{\rightarrow}$  10  $\stackrel{\text{out}}{\rightarrow}$  00) with  $\Gamma_{10}\Delta_t=1$ . The final polarization (teal) is responsible for the emitted signal, which is then passed through a frequency bandpass filter to emulate monochromator detection (Fig. 4.2c). The resulting signal is explored in 2D delay space (Fig. 4.2d), 2D frequency space (Fig. 4.2f), and hybrid delay-frequency space (Fig. 4.2e). The detuning frequency axes are also normalized by the pulse bandwidth,  $\Delta_\omega$ .

We now consider the generalized Liouville pathway  $L: \rho_0 \xrightarrow{x} \rho_1 \xrightarrow{y} \rho_2 \xrightarrow{z} \rho_3 \xrightarrow{\text{out}} \rho_4$ , where x, y, and z denote properties of the first, second, and third pulse, respectively, and indices 0, 1, 2, 3, and 4 define the properties of the ground state, first, second, third, and fourth density matrix elements, respectively. Fig. 4.2b demonstrates the correspondence between x, y, z notation and 1, 2, 2' notation for the laser pulses with pathway  $V\gamma$ .<sup>2</sup>

The electric field emitted from a Liouville pathway is proportional to the polarization created by the third-order coherence:

$$E_L(t) = i\mu_4 \rho_3(t). \tag{4.6}$$

Equation 4.6 assumes perfect phase-matching and no pulse distortions through propagation. Equation

<sup>&</sup>lt;sup>2</sup>For elucidation of the relationship between the generalized Liouville pathway notation and the specific parameters for each Liouville pathway, see Table S1 in the Supplementary Information.

4.5 shows that the output field for this Liouville pathway is

$$E_{L}(t) = \frac{i}{8} \lambda_{1} \lambda_{2} \lambda_{3} \mu_{1} \mu_{2} \mu_{3} \mu_{4} e^{i(\kappa_{1} \omega_{x} \tau_{x} + \kappa_{2} \omega_{y} \tau_{y} + \kappa_{3} \omega_{z} \tau_{z})} e^{-i(\kappa_{3} \omega_{z} + \kappa_{2} \omega_{y} + \kappa_{1} \omega_{x})t}$$

$$\times \iiint_{-\infty}^{\infty} c_{z}(t - u - \tau_{z}) c_{y}(t - u - v - \tau_{y}) c_{x}(t - u - v - w - \tau_{x}) R_{L}(u, v, w) dw \ dv \ du,$$

$$(4.7)$$

$$R_L(u, v, w) = \Theta(w)e^{-(\Gamma_1 + i\kappa_1\Omega_{1x})w}\Theta(v)e^{-(\Gamma_2 + i[\kappa_1\Omega_{1x} + \kappa_2\Omega_{2y}])v}\Theta(u)e^{-(\Gamma_3 + i[\kappa_1\Omega_{1x} + \kappa_2\Omega_{2y} + \kappa_3\Omega_{3z}])u}, \quad (4.8)$$

where  $R_L$  is the third-order response function for the Liouville pathway. The total electric field will be the superposition of all the Liouville pathways:

$$E_{\text{tot}} = \sum_{l} E_{L}(t). \tag{4.9}$$

For the superposition of Equation 4.9 to be non-canceling, certain symmetries between the pathways must be broken. In general, this requires one or more of the following inequalities:  $\Gamma_{10} \neq \Gamma_{21}$ ,  $\omega_{10} \neq \omega_{21}$ , and/or  $\sqrt{2}\mu_{10} \neq \mu_{21}$ . Our simulations use the last inequality, which is important in two-level systems  $(\mu_{21}=0)$  and in systems where state-filling dominates the non-linear response, such as in semiconductor excitons. The exact ratio between  $\mu_{10}$  and  $\mu_{21}$  affects the absolute amplitude of the field, but does not affect the multidimensional line shape. Importantly, the dipole inequality does not break the symmetry of double quantum coherence pathways (time-orderings II and IV), so such pathways are not present in our analysis.

In MR-CMDS, a monochromator resolves the driven output frequency from other nonlinear output frequencies, which in our case is  $\omega_m = \omega_1 - \omega_2 + \omega_{2'} = \omega_1$ . The monochromator can also enhance spectral resolution, as we show in Section 4.4.1. In this simulation, the detection is emulated by transforming  $E_{\text{tot}}(t)$  into the frequency domain, applying a narrow bandpass filter,  $M(\omega)$ , about  $\omega_1$ , and applying amplitude-scaled detection:

$$S_{\text{tot}}(\omega_1, \omega_2, \tau_{21}, \tau_{22'}) = \sqrt{\int |M(\omega - \omega_1)E_{\text{tot}}(\omega)|^2 d\omega}, \tag{4.10}$$

where  $E_{\text{tot}}(\omega)$  denotes the Fourier transform of  $E_{\text{tot}}(t)$  (see Fig. 4.2c). For M we used a rectangular

function of width  $0.408\Delta_{\omega}$ . The arguments of  $S_{\rm tot}$  refer to the *experimental* degrees of freedom. The signal delay dependence is parameterized with the relative delays  $\tau_{21}$  and  $\tau_{22'}$ , where  $\tau_{nm} = \tau_n - \tau_m$  (see Fig. 4.2b). Table S1 summarizes the arguments for each Liouville pathway. Fig. 4.2f shows the 2D  $(\omega_1, \omega_2)$   $S_{\rm tot}$  spectrum resulting from the pulse delay times represented in Fig. 4.2b.

## 4.2.1 Inhomogeneity

Inhomogeneity is isolated in CMDS through both spectral signatures, such as line-narrowing [Besemann2004, Oudar1980, Carlson1990, Riebe1988, Steehler1985], and temporal signatures, such as photon echoes [Weiner1985, Agarwal2002]. We simulate the effects of static inhomogeneous broadening by convolving the homogeneous response with a Gaussian distribution function. Further details of the convolution are in Appendix 4.3.2. Dynamic broadening effects such as spectral diffusion are beyond the scope of this work.

# 4.3 Methods

A matrix representation of differential equations of the type in Equation 4.7 was numerically integrated for parallel computation of Liouville elements (see SI for details).[Dick1983, Gelin2005] The lower bound of integration was  $2\Delta_t$  before the first pulse, and the upper bound was  $5\Gamma_{10}^{-1}$  after the last pulse, with step sizes much shorter than the pulse durations. Integration was performed in the FID rotating frame; the time steps were chosen so that both the system-pulse difference frequencies and the pulse envelope were well-sampled.

The following simulations explore the four-dimensional  $(\omega_1, \omega_2, \tau_{21}, \tau_{22'})$  variable space. Both frequencies are scanned about the resonance, and both delays are scanned about pulse overlap. We explored the role of sample dephasing rate by calculating signal for systems with dephasing rates such that  $\Gamma_{10}\Delta_t=0.5,1$ , and 2. Inhomogeneous broadening used a spectral FWHM,  $\Delta_{\rm inhom}$ , that satisfied  $\Delta_{\rm inhom}/\Delta_\omega=0,0.5,1$ , and 2 for the three dephasing rates. For all these dimensions, both  $\rho_3(t;\omega_1,\omega_2,\tau_{21},\tau_{22'})$  and  $S_{\rm tot}(\omega_1,\omega_2,\tau_{21},\tau_{22'})$  are recorded for each unique Liouville pathway. Our simulations were done using

the open-source SciPy library.[Oliphant2007]

#### 4.3.1 Characteristics of Driven and Impulsive Response

The changes in the spectral line shapes described in this work are best understood by examining the driven/continuous wave (CW) and impulsive limits of Equations 4.5 and 4.7. The driven limit is achieved when pulse durations are much longer than the response function dynamics:  $\Delta_t |\Gamma_f + i\kappa_f \Omega_{fx}| \gg 1$ . In this limit, the system will adopt a steady state over excitation:  $d\rho/dt \approx 0$ . Neglecting phase factors, the driven solution to Equation 4.5 will be

$$\tilde{\rho}_f(t) = \frac{\lambda_f \mu_f}{2} \frac{c_x(t - \tau_x) e^{i\kappa_f \Omega_{fx} t}}{\kappa_f \Omega_{fx}} \tilde{\rho}_i(t). \tag{4.11}$$

The frequency and temporal envelope of the excitation pulse controls the coherence time evolution, and the relative amplitude and phase of the coherence is directly related to detuning from resonance.

The impulsive limit is achieved when the excitation pulses are much shorter than response function dynamics:  $\Delta_t |\Gamma_f + i\kappa_f \Omega_{fx}| \ll 1$ . The full description of the temporal evolution has two separate expressions: one for times when the pulse is interacting with the system, and one for times after pulse interaction. Both expressions are important when describing CMDS experiments.

For times after the pulse interaction,  $t \gtrsim \tau_x + \Delta_t$ , the field-matter coupling is negligible. The evolution for these times, on resonance, is given by

$$\tilde{\rho}_f(t) = \frac{i\lambda_f \mu_f}{2} \tilde{\rho}_i(\tau_x) \int c_x(u) du \ e^{-\Gamma_f(t-\tau_x)}. \tag{4.12}$$

This is classic free induction decay (FID) evolution: the system evolves at its natural frequency and decays at rate  $\Gamma_f$ . It is important to note that, while this expression is explicitly derived from the impulsive limit, FID behavior is not exclusive to impulsive excitation, as we have defined it. A latent FID will form if the pulse vanishes at a fast rate relative to the system dynamics.

For evaluating times near pulse excitation,  $t \lesssim au_{x} + \Delta_{t}$ , we implement a Taylor expansion in the response

function about zero:  $e^{-(\Gamma_f + i\kappa_f\Omega_{fx})u} = 1 - (\Gamma_f + i\kappa_f\Omega_{fx})u + \cdots$ . Our impulsive criterion requires that a low order expansion will suffice; it is instructive to consider the result of the first order expansion of Equation 4.5:

$$\tilde{\rho}_{f}(t) = \frac{i\lambda_{f}\mu_{f}}{2} e^{-i\kappa_{f}\omega_{x}\tau_{x}} e^{-i\kappa_{f}\Omega_{fx}t} \tilde{\rho}_{i}(\tau_{x})$$

$$\times \left[ (1 - (\Gamma_{f} + i\kappa_{f}\Omega_{fx})(t - \tau_{x})) \int_{-\infty}^{t - \tau_{x}} c_{x}(u)du \right]$$

$$+ (\Gamma_{f} + i\kappa_{f}\Omega_{fx}) \int_{-\infty}^{t - \tau_{x}} c_{x}(u)u \ du$$
(4.13)

During this time  $\tilde{\rho}_f$  builds up roughly according to the integration of the pulse envelope. The build-up is integrated because the pulse transfers energy before appreciable dephasing or detuning occurs. Contrary to the expectation of impulsive evolution, the evolution of  $\tilde{\rho}_f$  is explicitly affected by the pulse frequency, and the temporal profile evolves according to the pulse.

It is important to recognize that the impulsive limit is defined not only by having slow relaxation relative to the pulse duration, but also by small detuning relative to the pulse bandwidth (as is stated in the inequality). As detuning increases, the higher orders of the response function Taylor expansion will be needed to describe the rise time, and the driven limit of Equation 4.11 will become valid. The details of this build-up time can often be neglected in impulsive approximations because build-up contributions are often negligible in analysis; the period over which the initial excitation occurs is small in comparison to the free evolution of the system. The build-up behavior can be emphasized by the measurement, which makes Equation 4.13 important.

We now consider full Liouville pathways in the impulsive and driven limits of Equation 4.7. For the driven limit, Equation 4.7 can be reduced to

$$E_{L}(t) = \frac{1}{8} \lambda_{1} \lambda_{2} \lambda_{3} \mu_{1} \mu_{2} \mu_{3} \mu_{4} e^{-i(\kappa_{1} \omega_{x} \tau_{x} + \kappa_{2} \omega_{y} \tau_{y} + \kappa_{3} \omega_{z} \tau_{z})}$$

$$\times e^{i(\kappa_{3} \omega_{z} + \kappa_{2} \omega_{y} + \kappa_{1} \omega_{x})t}$$

$$\times c_{z}(t - \tau_{z}) c_{y}(t - \tau_{y}) c_{x}(t - \tau_{x})$$

$$\times \frac{1}{\kappa_{1} \Omega_{1x} - i\Gamma_{1}} \frac{1}{\kappa_{1} \Omega_{1x} + \kappa_{2} \Omega_{2y} - i\Gamma_{2}}$$

$$\times \frac{1}{\kappa_{1} \Omega_{1y} + \kappa_{2} \Omega_{2y} + \kappa_{3} \Omega_{3z} - i\Gamma_{3}}.$$

$$(4.14)$$

It is important to note that the signal depends on the multiplication of all the fields; pathway discrimination based on pulse time-ordering is not achievable because polarizations exists only when all pulses are overlapped. This limit is the basis for frequency-domain techniques. Frequency axes, however, are not independent because the system is forced to the laser frequency and influences the resonance criterion for subsequent excitations. As an example, observe that the first two resonant terms in Equation 4.14 are maximized when  $\omega_x = |\omega_1|$  and  $\omega_y = |\omega_2|$ . If  $\omega_x$  is detuned by some value  $\varepsilon$ , however, the occurrence of the second resonance shifts to  $\omega_y = |\omega_2| + \varepsilon$ , effectively compensating for the  $\omega_x$  detuning. This shifting of the resonance results in 2D line shape correlations.

If the pulses do not temporally overlap  $(\tau_x + \Delta_t \lesssim \tau_y + \Delta_t \lesssim \tau_z + \Delta_t \lesssim t)$ , then the impulsive solution to the full Liouville pathway of Equation 4.7 is

$$E_{L}(t) = \frac{i}{8} \lambda_{1} \lambda_{2} \lambda_{3} \mu_{1} \mu_{2} \mu_{3} \mu_{4} e^{i(\omega_{1} + \omega_{2} + \omega_{3})t}$$

$$\times \int c_{x}(w) dw \int c_{y}(v) dv \int c_{z}(u) du$$

$$\times e^{-\Gamma_{1}(\tau_{y} - \tau_{x})} e^{-\Gamma_{2}(\tau_{z} - \tau_{y})} e^{-\Gamma(t - \tau_{z})}.$$

$$(4.15)$$

Pathway discrimination is demonstrated here because the signal is sensitive to the time-ordering of the pulses. This limit is suited for delay scanning techniques. The emitted signal frequency is determined by the system and can be resolved by scanning a monochromator.

The driven and impulsive limits can qualitatively describe our simulated signals at certain frequency and delay combinations. Of the three expressions, the FID limit most resembles signal when pulses are near resonance and well-separated in time (so that build-up behavior is negligible). The build-up limit approximates well when pulses are near-resonant and arrive together (so that build-up behavior is emphasized). The driven limit holds for large detunings, regardless of delay.

#### 4.3.2 Convolution Technique for Inhomogeneous Broadening




Figure 4.3: Overview of the convolution. (a) The homogeneous line shape. (b) The distribution function, K, mapped onto laser coordinates. (c) The resulting ensemble line shape computed from the convolution. The thick black line represents the FWHM of the distribution function.

Here we describe how to transform the data of a single reference oscillator signal to that of an inhomogeneous distribution. The oscillators in the distribution are allowed have arbitrary energies for their states, which will cause frequency shifts in the resonances. To show this, we start with a modified, but equivalent, form of Equation 4.4:

$$\frac{d\tilde{\rho}_f}{dt} = -\Gamma_f \tilde{\rho}_f + \frac{i}{2} \lambda_f \mu_f c_x (t - \tau_x) 
\times e^{i\kappa_f (\vec{k} \cdot z + \omega_x \tau_x)} e^{-i\kappa_f (\omega_x - |\omega_f|)t} \tilde{\rho}_i(t).$$
(4.16)

We consider two oscillators with transition frequencies  $\omega_f$  and  $\omega_f' = \omega_f + \delta$ . So long as  $|\delta| \leq \omega_f$  (so that  $|\omega_f + \delta| = |\omega_f| + \delta$  and thus the rotating wave approximation does not change), Equation 4.16 shows that the two are related by

$$\frac{d\tilde{\rho}_f'}{dt}(t;\omega_x) = \frac{d\tilde{\rho}_f}{dt}(t;\omega_x - \delta)e^{i\kappa_f\delta\tau_x}.$$
(4.17)

Because both coherences are assumed to have the same initial conditions  $(\rho_0(-\infty) = \rho_0'(-\infty) = 0)$ , the equality also holds when both sides of the equation are integrated. The phase factor  $e^{i\kappa_f\delta\tau_x}$  in the substitution arises from Equation 4.1, where the pulse carrier frequency maintains its phase within the pulse envelope for all delays.

The resonance translation can be extended to higher order signals as well. For a third-order signal, we compare systems with transition frequencies  $\omega'_{10}=\omega_{10}+a$  and  $\omega'_{21}=\omega_{21}+b$ . The extension of Equation 4.17 to pathway  $V\beta$  gives

$$\tilde{\rho}_{3}'(t;\omega_{2},\omega_{2}',\omega_{1}) = \tilde{\rho}_{3}(t;\omega_{2}-a,\omega_{2'}-a,\omega_{1}-b) \times e^{i\kappa_{2}a\tau_{2}}e^{i\kappa_{2'}a\tau_{2'}}e^{i\kappa_{1}b\tau_{1}}.$$
(4.18)

The translation of each laser coordinate depends on which transition is made (e.g. a for transitions between  $|0\rangle$  and  $|1\rangle$  or b for transitions between  $|1\rangle$  and  $|2\rangle$ ), so the exact translation relation differs between pathways. We can now compute the ensemble average of signal for pathway  $V\beta$  as a convolution

between the distribution function of the system, K(a, b), and the single oscillator response:

$$\langle \tilde{\rho}_{3}(t; \omega_{2}, \omega_{2'}, \omega_{1}) \rangle = \iint K(a, b)$$

$$\times \tilde{\rho}_{3}(t; \omega_{2} + a, \omega_{2'} + a, \omega_{1} + b)$$

$$\times e^{i\kappa_{2}a\tau_{2}} e^{i\kappa_{2'}a\tau_{2'}} e^{i\kappa_{1}b\tau_{1}} da \ db.$$

$$(4.19)$$

For this work, we restrict ourselves to a simpler ensemble where all oscillators have equally spaced levels (i.e. a = b). This makes the translation identical for all pathways and reduces the dimensionality of the convolution. Since pathways follow the same convolution we may also perform the convolution on the total signal field:

$$\langle E_{\text{tot}}(t) \rangle = \sum_{L} \mu_{4,L} \int K(a,a)$$

$$\times \tilde{\rho}_{3,L}(t; \omega_{x} - a, \omega_{y} - a\omega_{z} - a)$$

$$\times e^{ia(\kappa_{x}\tau_{x} + \kappa_{y}\tau_{y} + \kappa_{z}\tau_{z})} da.$$

$$(4.20)$$

Furthermore, since  $\kappa=-1$  for  $E_1$  and  $E_{2'}$ , while  $\kappa=1$  for  $E_2$ , we have  $e^{ia(\kappa_{\chi}\tau_{\chi}+\kappa_{y}\tau_{y}+\kappa_{z}\tau_{z})}=e^{-ia(\tau_{1}-\tau_{2}+\tau_{2'})}$  for all pathways. Equivalently, if the electric field is parameterized in terms of laser coordinates  $\omega_1$  and  $\omega_2$ , the ensemble field can be calculated as

$$\langle E_{\text{tot}}(t;\omega_1,\omega_2)\rangle = \int K(a,a)E_{\text{tot}}(t;\omega_1-a,\omega_2-a) \times e^{-ia(\tau_1-\tau_2+\tau_{2'})}da.$$
(4.21)

which is a 1D convolution along the diagonal axis in frequency space. Fig. 4.3 demonstrates the use of Equation 4.21 on a homogeneous line shape.

# 4.4 Results

We now present portions of our simulated data that highlight the dependence of the spectral line shapes and transients on excitation pulse width, the dephasing rate, detuning from resonance, the pulse delay times, and inhomogeneous broadening.

# 4.4.1 Evolution of single coherence




Figure 4.4: The relative importance of FID and driven response for a single quantum coherence as a function of the relative dephasing rate (values of  $\Gamma_{10}\Delta_t$  are shown inset). The black line shows the coherence amplitude profile, while the shaded color indicates the instantaneous frequency (see colorbar). For all cases, the pulsed excitation field (gray line, shown as electric field amplitude) is slightly detuned (relative detuning,  $\Omega_{\rm fx}/\Delta_\omega=0.1$ ).

It is illustrative to first consider the evolution of single coherences,  $\rho_0 \xrightarrow{\times} \rho_1$ , under various excitation conditions. Fig. 4.4 shows the temporal evolution of  $\rho_1$  with various dephasing rates under Gaussian excitation. The value of  $\rho_1$  differs only by phase factors between various Liouville pathways (this can be verified by inspection of Equation 4.5 under the various conditions in Table S1), so the profiles in Fig. 4.4 apply for the first interaction of any pathway. The pulse frequency was detuned from resonance so that frequency changes could be visualized by the color bar, but the detuning was kept slight so that it did not appreciably change the dimensionless product,  $\Delta_t \left(\Gamma_f + i\kappa_f \Omega_{fx}\right) \approx \Gamma_{10} \Delta_t$ . In this case, the evolution demonstrates the maximum impulsive character the transient can achieve. The instantaneous frequency,  $d\varphi/dt$ , is defined as

$$\frac{d\varphi}{dt} = \frac{d}{dt} \tan^{-1} \left( \frac{\operatorname{Im} (\rho_1(t))}{\operatorname{Re} (\rho_1(t))} \right). \tag{4.22}$$

The cases of  $\Gamma_{10}\Delta_t=0(\infty)$  agree with the impulsive (driven) expressions derived in Appendix 4.3.1. For  $\Gamma_{10}\Delta_t=0$ , the signal rises as the integral of the pulse and has instantaneous frequency close to that of the pulse (Equation 4.13), but as the pulse vanishes, the signal adopts the natural system frequency and decay rate (Equation 4.12). For  $\Gamma_{10}\Delta_t=\infty$ , the signal follows the amplitude and frequency of the pulse for all times (the driven limit, Equation 4.11).

The other three cases show a smooth interpolation between limits. As  $\Gamma_{10}\Delta_t$  increases from the impulsive limit, the coherence within the pulse region conforms less to a pulse integral profile and more to a pulse envelope profile. In accordance, the FID component after the pulse becomes less prominent, and the instantaneous frequency pins to the driving frequency more strongly through the course of evolution. The trends can be understood by considering the differential form of evolution (Equation 4.4), and the time-dependent balance of optical coupling and system relaxation. We note that our choices of  $\Gamma_{10}\Delta_t=2.0,1.0$ , and 0.5 give coherences that have mainly driven, roughly equal driven and FID parts, and mainly FID components, respectively. FID character is difficult to isolate when  $\Gamma_{10}\Delta_t=2.0$ .




Figure 4.5: Pulsed excitation of a single quantum coherence and its dependence on the pulse detuning. In all cases, relative dephasing is kept at  $\Gamma_{10}\Delta_t=1$ . (a) The relative importance of FID and driven response for a single quantum coherence as a function of the detuning (values of relative detuning,  $\Omega_{fx}/\Delta_{\omega}$ , are shown inset). The color indicates the instantaneous frequency (scale bar on right), while the black line shows the amplitude profile. The gray line is the electric field amplitude. (b) The time-integrated coherence amplitude as a function of the detuning. The integrated amplitude is collected both with (teal) and without (magenta) a tracking monochromator that isolates the driven frequency components. For comparison, the Green's function of the single quantum coherence is also shown (amplitude is black, hashed; imaginary is black, solid). In all plots, the gray line is the electric field amplitude.

Fig. 4.5a shows the temporal evolution of  $\rho_1$  at several values of  $\Omega_{1x}/\Delta_{\omega}$  with  $\Gamma_{10}\Delta_t=1.^3$  As detuning increases, total amplitude decreases, FID character vanishes, and  $\rho_1$  assumes a more driven character, as expected. During the excitation,  $\rho_1$  maintains a phase relationship with the input field (as seen by the instantaneous frequency in Fig. 4.5a). The coherence will persist beyond the pulse duration only if the pulse transfers energy into the system; FID evolution equates to absorption. The FID is therefore sensitive to the absorptive (imaginary) line shape of a transition, while the driven response is the composite of both absorptive and dispersive components. If the experiment isolates the latent FID response, there is consequently a narrower spectral response. This spectral narrowing can be seen in Fig. 4.5a by comparing the coherence amplitudes at t=0 (driven) and at  $t/\Delta_t=2$  (FID); amplitudes for all  $\Omega_{fx}/\Delta_{\omega}$  values shown are comparable at t=0, but the lack of FID formation for  $\Omega_{fx}/\Delta_{\omega}=\pm 2$  manifests as a visibly disproportionate amplitude decay.<sup>4</sup> Many ultrafast spectroscopies take advantage of the latent FID to suppress non-resonant background, improving signal to noise.[Lagutchev2007, Lagutchev2010, Donaldson2010, Donaldson2008]

In driven experiments, the output frequency and line shape are fully constrained by the excitation beams. In such experiments, there is no additional information to be resolved in the output spectrum. The situation changes in the mixed domain, where  $E_{\rm tot}$  contains FID signal that lasts longer than the pulse duration. Fig. 4.5a provides insight on how frequency-resolved detection of coherent output can enhance resolution when pulses are spectrally broad. Without frequency-resolved detection, mixed-domain resonance enhancement occurs in two ways: (1) the peak amplitude increases, and (2) the coherence duration increases due to the FID transient. Frequency-resolved detection can further discriminate against detuning by requiring that the driving frequency agrees with latent FID. The implications of discrimination are most easily seen in Fig. 4.5a with  $\Omega_{1x}/\Delta_{\omega}=\pm 1$ , where the system frequency moves from the driving frequency to the FID frequency. When the excitation pulse frequency is scanned, the resonance will be more sensitive to detuning by isolating the driven frequency (tracking the monochromator with the excitation source).

The functional form of the measured line shape can be deduced by considering the frequency domain

<sup>&</sup>lt;sup>3</sup> See Supplementary Fig. S3 for a Fourier domain representation of Fig. 4.5a.

<sup>&</sup>lt;sup>4</sup> See Supplementary Fig. S4 for explicit plots of  $\rho_1(\Omega_{fx}/\Delta_{\omega})$  at discrete  $t/\Delta_t$  values.

form of Equation 4.5 (assume  $\rho_i=1$  and  $au_{\mathsf{x}}=0$ ):

$$\tilde{\rho}_f(\omega) = \frac{i\lambda_f \mu_f}{2\sqrt{2\pi}} \cdot \frac{\mathcal{F}\left\{c_{\mathsf{x}}\right\} \left(\omega - \kappa_f \Omega_{f\mathsf{x}}\right)}{\Gamma_f + i\omega},\tag{4.23}$$

where  $\mathcal{F}\left\{c_{\mathsf{x}}\right\}\left(\omega\right)$  denotes the Fourier transform of  $c_{\mathsf{x}}$ , which in our case gives

$$\mathcal{F}\left\{c_{x}\right\}\left(\omega\right) = \frac{1}{\sqrt{2\pi}}e^{-\frac{\left(\Delta_{t}\omega\right)^{2}}{4\ln2}}.\tag{4.24}$$

For squared-law detection of  $\rho_f$ , the importance of the tracking monochromator is highlighted by two limits of Equation 4.23:

- When the transient is not frequency resolved, sig  $\approx \int |\tilde{\rho}_f(\omega)|^2 d\omega$  and the measured line shape will be the convolution of the pulse envelope and the intrinsic (Green's function) response (Fig. 4.5b, magenta).
- When the driven frequency is isolated, sig  $\approx |\tilde{\rho}_f(\kappa_f\Omega_{fx})|^2$  and the measured line shape will give the un-broadened Green's function (Fig. 4.5b, teal).

Monochromatic detection can remove broadening effects due to the pulse bandwidth. For large  $\Gamma_{10}\Delta_t$  values, FID evolution is negligible at all  $\Omega_{fx}/\Delta_\omega$  values and the monochromator is not useful. Fig. 4.5b shows the various detection methods for the relative dephasing rate of  $\Gamma_{10}\Delta_t=1$ .

### 4.4.2 Evolution of single Liouville pathway




Figure 4.6: Changes to the 2D frequency response of a single Liouville pathway ( $I\gamma$ ) at different delay values. The normalized dephasing rate is  $\Gamma_{10}\Delta_t=1$ . Left: The 2D delay response of pathway  $I\gamma$  at triple resonance. Right: The 2D frequency response of pathway  $I\gamma$  at different delay values. The delays at which the 2D frequency plots are collected are indicated on the delay plot; compare 2D spectrum frame color with dot color on 2D delay plot.

We now consider the multidimensional response of a single Liouville pathway involving three pulse interactions. In a multi-pulse experiment,  $\rho_1$  acts as a source term for  $\rho_2$  (and subsequent excitations). The spectral and temporal features of  $\rho_1$  that are transferred to  $\rho_2$  depend on when the subsequent pulse arrives. Time-gating later in  $\rho_1$  evolution will produce responses with FID behavior, while time-gating  $\rho_1$  in the presence of the initial pulse will produce driven responses. An analogous relationship holds for  $\rho_3$  with its source term  $\rho_2$ . As discussed above, signal that time-gates FID evolution gives narrower spectra than driven-gated signal. As a result, the spectra of even single Liouville pathways will change based on pulse delays.

The final coherence will also be frequency-gated by the monochromator. The monochromator isolates signal at the fully driven frequency  $\omega_{\text{out}} = \omega_1$ . The monochromator will induce line-narrowing to the extent that FID takes place. It effectively enforces a frequency constraint that acts as an additional resonance condition,  $\omega_{\text{out}} = \omega_1$ . The driven frequency will be  $\omega_1$  if  $E_1$  is the last pulse interaction (time-orderings V and VI), and the monochromator tracks the coherence frequency effectively. If  $E_1$  is not the last interaction, the output frequency may not be equal to the driven frequency, and the monochromator plays a more complex role.

We demonstrate this delay dependence using the multidimensional response of the  $I\gamma$  Liouville pathway as an example (see Fig. 4.1). Fig. 4.6 shows the resulting 2D delay profile of pathway  $I\gamma$  signals for  $\Gamma_{10}\Delta_t=1$  (left) and the corresponding  $\omega_1,\omega_2$  2D spectra at several pulse delay values (right). The spectral changes result from changes in the relative importance of driven and FID components. The prominence of FID signal can change the resonance conditions; Table 4.1 summarizes the changing resonance conditions for each of the four delay coordinates studied. Since  $E_1$  is not the last pulse in pathway  $I\gamma$ , the tracking monochromator must also be considered.<sup>5</sup>

<sup>&</sup>lt;sup>5</sup> See Supplementary Fig. S5 for a representation of Fig. 5 simulated without monochromator frequency filtering  $(M(\omega - \omega_1) = 1 \text{ for Equation 4.10}).$ 

Table 4.1: Conditions for peak intensity at different pulse delays for pathway I $\gamma$ . Delay Approximate Resonance Conditions

| Delay               |                      | Approximate Resonance Conditions |                                   |                                    |  |
|---------------------|----------------------|----------------------------------|-----------------------------------|------------------------------------|--|
| $	au_{21}/\Delta_t$ | $	au_{22'}/\Delta_t$ | $\rho_0 \xrightarrow{1} \rho_1$  | $ \rho_1 \xrightarrow{2} \rho_2 $ | $ \rho_2 \xrightarrow{2'} \rho_3 $ | $ ho_3  ightarrow 	ext{detection at } \omega_m = \omega_1$ |
| 0                   | 0                    | $\omega_1 = \omega_{10}$         | $\omega_1 = \omega_2$             | $\omega_1 = \omega_{10}$           | _  |
| 0                   | -2.4                 | $\omega_1 = \omega_{10}$         | $\omega_1 = \omega_2$             | $\omega_2 = \omega_{10}$           | $\omega_1=\omega_2$  |
| 2.4                 | 0                    | $\omega_1 = \omega_{10}$         | $\omega_2 = \omega_{10}$          | _                                  | $\omega_1=\omega_{10}$                                     |
| 2.4                 | -2.4                 | $\omega_1 = \omega_{10}$         | $\omega_2 = \omega_{10}$          | $\omega_2 = \omega_{10}$           | $\omega_1=\omega_2$  |

When the pulses are all overlapped ( $\tau_{21}=\tau_{22'}=0$ , lower right, orange), all transitions in the Liouville pathway are simultaneously driven by the incident fields. This spectrum strongly resembles the driven limit spectrum. For this time-ordering, the first, second, and third density matrix elements have driven resonance conditions of  $\omega_1=\omega_{10},\ \omega_1-\omega_2=0$ , and  $\omega_1-\omega_2+\omega_{2'}=\omega_{10}$ , respectively. The second resonance condition causes elongation along the diagonal, and since  $\omega_2=\omega_{2'}$ , the first and third resonance conditions are identical, effectively making  $\omega_1$  doubly resonant at  $\omega_{10}$  and resulting in the vertical elongation along  $\omega_1=\omega_{10}$ .

The other three spectra of Fig. 4.6 separate the pulse sequence over time so that not all interactions are driven. At  $\tau_{21}=0$ ,  $\tau_{22'}=-2.4\Delta_t$  (lower left, pink), the first two resonances remain the same as at pulse overlap (orange) but the last resonance is different. The final pulse,  $E_{2'}$ , is latent and probes  $\rho_2$  during its FID evolution after memory of the driven frequency is lost. There are two important consequences. Firstly, the third driven resonance condition is now approximated by  $\omega_{2'}=\omega_{10}$ , which makes  $\omega_1$  only singly resonant at  $\omega_1=\omega_{10}$ . Secondly, the driven portion of the signal frequency is determined only by the latent pulse:  $\omega_{\text{out}}=\omega_{2'}$ . Since our monochromator gates  $\omega_1$ , we have the detection-induced correlation  $\omega_1=\omega_{2'}$ . The net result is double resonance along  $\omega_1=\omega_2$ , and the vertical elongation of pulse overlap is strongly attenuated.

At  $\tau_{21}=2.4\Delta_t$ ,  $\tau_{22'}=0$  (upper right, purple), the first pulse  $E_1$  precedes the latter two, which makes the two resonance conditions for the input fields  $\omega_1=\omega_{10}$  and  $\omega_2=\omega_{10}$ . The signal depends on the FID conversion of  $\rho_1$ , which gives vertical elongation at  $\omega_1=\omega_{10}$ . Furthermore,  $\rho_1$  has no memory of  $\omega_1$  when  $E_2$  interacts, which has two important implications. First, this means the second resonance condition  $\omega_1=\omega_2$  and the associated diagonal elongation is now absent. Second, the final output polarization frequency content is no longer functional of  $\omega_1$ . Coupled with the fact that  $E_2$  and  $E_{2'}$  are coincident, so that the final coherence can be approximated as driven by these two, we can approximate the final frequency as  $\omega_{\rm out}=\omega_{10}-\omega_2+\omega_{2'}=\omega_{10}$ . Surprisingly, the frequency content of the output is strongly independent of all pulse frequencies. The monochromator narrows the  $\omega_1=\omega_{10}$  resonance. The  $\omega_1=\omega_{10}$  resonance condition now depends on the monochromator slit width, the FID propagation of  $\rho_1$ , the spectral bandwidth of  $\rho_3$ ; its spectral width is not easily related to material parameters. This resonance demonstrates the importance of the detection scheme for experiments and how the optimal

detection can change depending on the pulse delay time.

Finally, when all pulses are well-separated ( $\tau_{21}=-\tau_{22'}=2.4\Delta_t$ , upper left, cyan), each resonance condition is independent and both  $E_1$  and  $E_2$  require FID buildup to produce final output. The resulting line shape is narrow in all directions. Again, the emitted frequency does not depend on  $\omega_1$ , yet the monochromator resolves the final coherence at frequency  $\omega_1$ . Since the driven part of the final interaction comes from  $E_{2'}$ , and since the monochromator track  $\omega_1$ , the output signal will increase when  $\omega_1=\omega_{2'}$ . As a result, the line shape acquires a diagonal character.

The changes in line shape seen in Fig. 4.6 have significant ramifications for the interpretations and strategies of MR-CMDS in the mixed domain. Time-gating has been used to isolate the 2D spectra of a certain time-ordering[Meyer2004, Pakoulev2006, Donaldson2007], but here we show that time-gating itself causes significant line shape changes to the isolated pathways. The phenomenon of time-gating can cause frequency and delay axes to become functional of each other in unexpected ways.

#### 4.4.3 Temporal pathway discrimination




Figure 4.7: Comparison of the 2D delay response for different relative dephasing rates (labeled atop each column). All pulses are tuned to exact resonance. In each 2D delay plot, the signal amplitude is depicted by the colors. The black contour lines show signal purity, P (see Equation 4.25), with purity values denoted on each contour. The small plots above each 2D delay plot examine a  $\tau_{22'}$  slice of the delay response ( $\tau_{21}=0$ ). The plot shows the total signal (black), as well as the component time-orderings VI (orange), V (purple), and III or I (teal).

In the last section we showed how a single pathway's spectra can evolve with delay due to pulse effects and time gating. In real experiments, evolution with delay is further complicated by the six time-orderings/sixteen pathways present in our three-beam experiment (see Fig. 4.1). Each time-ordering has different resonance conditions. When signal is collected near pulse overlap, multiple time-orderings contribute. To identify these effects, we start by considering how strongly time-orderings are isolated at each delay coordinate.

While the general idea of using time delays to enhance certain time-ordered regions is widely applied, quantitation of this discrimination is rarely explored. Because the temporal profile of the signal is dependent on both the excitation pulse profile and the decay dynamics of the coherence itself, quantitation of pathway discrimination requires simulation.

Fig. 4.7 shows the 2D delay space with all pathways present for  $\omega_1, \omega_2 = \omega_{10}$ . It illustrates the interplay of pulse width and system decay rates on the isolation of time-ordered pathways. The color bar shows the signal amplitude. Signal is symmetric about the  $\tau_{21} = \tau_{22'}$  line because when  $\omega_1 = \omega_2$ ,  $E_1$  and  $E_{2'}$  interactions are interchangeable:  $S_{\text{tot}}(\tau_{21}, \tau_{22'}) = S_{\text{tot}}(\tau_{22'}, \tau_{21})$ . The overlaid black contours represent signal "purity," P, defined as the relative amount of signal that comes from the dominant pathway at that delay value:

$$P(\tau_{21}, \tau_{22'}) = \frac{\max \{S_L(\tau_{21}, \tau_{22'})\}}{\sum_L S_L(\tau_{21}, \tau_{22'})}.$$
(4.25)

The dominant pathway (max  $\{S_L(\tau_{21}, \tau_{22'})\}$ ) at given delays can be inferred by the time-ordered region defined in Fig. 4.2d. The contours of purity generally run parallel to the time-ordering boundaries with the exception of time-ordered regions II and IV, which involve the double quantum coherences that have been neglected.

A commonly-employed metric for temporal selectivity is how definitively the pulses are ordered. This metric agrees with our simulations. The purity contours have a weak dependence on  $\Delta_t \Gamma_{10}$  for  $|\tau_{22'}|/\Delta_t < 1$  or  $|\tau_{21}|/\Delta_t < 1$  where there is significant pathway overlap and a stronger dependence at larger values where the pathways are well-isolated. Because responses decay exponentially, while pulses decay as Gaussians, there always exist delays where temporal discrimination is possible. As  $\Delta_t \Gamma_{10} \to \infty$ , however, such discrimination is only achieved at vanishing signal intensities; the contour of P=0.99 across

our systems highlights this trend.

# 4.4.4 Multidimensional line shape dependence on pulse delay time




Figure 4.8: Evolution of the 2D frequency response as a function of  $\tau_{21}$  (labeled inset) and the influence of the relative dephasing rate ( $\Gamma_{10}\Delta_t=0.5$  (red), 1.0 (green), and 2.0 (blue)). In all plots  $\tau_{22'}=0$ . To ease comparison between different dephasing rates, the colored line contours (showing the half-maximum) for all three relative dephasing rates are overlaid. The colored histograms below each 2D frequency plot show the relative weights of each time-ordering for each relative dephasing rate. Contributions from V and VI are grouped together because they have equal weights at  $\tau_{22'}=0$ .

In the previous sections we showed how pathway spectra and weights evolve with delay. This section ties the two concepts together by exploring the evolution of the spectral line shape over a span of  $\tau_{21}$  delay times that include all pathways. It is a common practice to explore spectral evolution against  $\tau_{21}$  because this delay axis shows population evolution in a manner analogous to pump-probe spectroscopies. The  $\vec{k}_2$  and  $\vec{k}_2$  interactions correspond to the pump, and the  $\vec{k}_1$  interaction corresponds to the probe. Time-orderings V and VI are the normal pump-probe time-orderings, time-ordering III is a mixed pump-probe-pump ordering (so-called pump polarization coupling), and time-ordering I is the probe-pump ordering (so-called perturbed FID). Scanning  $\tau_{21}$  through pulse overlap complicates interpretation of the line shape due to the changing nature and balance of the contributing time-orderings. At  $\tau_{21} > 0$ , time-ordering I dominates; at  $\tau_{21} = 0$ , all time-orderings contribute equally; at  $\tau_{21} < 0$  time-orderings V and VI dominate (Fig. 4.7). Conventional pump-probe techniques recognized these complications long ago,[BritoCruz1988, Palfrey1985] but the extension of these effects to MR-CMDS has not previously been done.

Fig. 4.8 shows the MR-CMDS spectra, as well as histograms of the pathway weights, while scanning  $\tau_{21}$  through pulse overlap. The colored histogram bars and line shape contours correspond to different values of the relative dephasing rate,  $\Gamma_{10}\Delta_t$ . The contour is the half-maximum of the line shape.<sup>6</sup> The dependence of the line shape amplitude on  $\tau_{21}$  can be inferred from Fig. 4.7.

The qualitative trend, as  $\tau_{21}$  goes from positive to negative delays, is a change from diagonal/compressed line shapes to much broader resonances with no correlation ( $\omega_1$  and  $\omega_2$  interact with independent resonances). Such spectral changes could be misinterpreted as spectral diffusion, where the line shape changes from correlated to uncorrelated as population time increases due to system dynamics. The system dynamics included here, however, contain no structure that would allow for such diffusion. Rather, the spectral changes reflect the changes in the majority pathway contribution, starting with time-ordering I pathways, proceeding to an equal admixture of I, III, V, and VI, and finishing at an equal balance of V and VI when  $E_1$  arrives well after  $E_2$  and  $E_{2'}$ . Time-orderings I and III both exhibit a spectral correlation in  $\omega_1$  and  $\omega_2$  when driven, but time-orderings V and VI do not. Moreover, such spectral correlation is forced near zero delay because the pulses time-gate the driven signals of the first

<sup>&</sup>lt;sup>6</sup>Supplementary Fig. S6 shows fully colored contour plots of each 2D frequency spectrum.

two induced polarizations. The monochromator detection also plays a dynamic role, because time-orderings V always VI always emit a signal at the monochromator frequency, while in time-orderings I and III the emitted frequency is not defined by  $\omega_1$ , as discussed above.

When we isolate time-orderings V and VI, we can maintain the proper scaling of FID bandwidth in the  $\omega_1$  direction because our monochromator can gate the final coherence. This gating is not possible in time-orderings I and III because the final coherence frequency is determined by  $\omega_{2'}$  which is identical to  $\omega_2$ .

There are differences in the line shapes for the different values of the relative dephasing rate,  $\Gamma_{10}\Delta_t$ . The spectral correlation at  $\tau_{21}/\Delta_t=2$  decreases in strength as  $\Gamma_{10}\Delta_t$  decreases. As we illustrated in Fig. 4.6, this spectral correlation is a signature of driven signal from temporal overlap of  $E_1$  and  $E_2$ ; the loss of spectral correlation reflects the increased prominence of FID in the first coherence as the field-matter interactions become more impulsive. This increased prominence of FID also reflects an increase in signal strength, as shown by  $\tau_{21}$  traces in Fig. 4.7. When all pulses are completely overlapped, ( $\tau_{21}=0$ ), each of the line shapes exhibit spectral correlation. At  $\tau_{21}/\Delta_t=-2$ , the line shape shrinks as  $\Gamma_{10}\Delta_t$  decreases, with the elongation direction changing from horizontal to vertical. The general shrinking reflects the narrowing homogeneous linewidth of the  $\omega_{10}$  resonance. In all cases, the horizontal line shape corresponds to the homogeneous linewidth because the narrow bandpass monochromator resolves the final  $\omega_1$  resonance. The change in elongation direction is due to the resolving power of  $\omega_2$ . At  $\Gamma_{10}\Delta_t=0.5$ , the resonance is broader than our pulse bandwidth and is fully resolved vertically. It is narrower than the  $\omega_1$  resonance because time-orderings V and VI interfere to isolate only the absorptive line shape along  $\omega_2$ . This narrowing, however, is unresolvable when the pulse bandwidth becomes broader than that of the resonance, which gives rise to a vertically elongated signal when  $\Gamma_{10}\Delta_t=0.5$ .




Figure 4.9: Transient  $(\omega_1)$  line shapes and their dependence on  $\omega_2$  frequency. The relative dephasing rate is  $\Gamma_{10}\Delta_t=1$  and  $\tau_{22'}=0$ . For each plot, the corresponding  $\omega_2$  value is shown as a light gray vertical line.

It is also common to represent data as "Wigner plots," where one axis is delay and the other is frequency.[Kohler2014, Aubock2012, Czech2015, Pakoulev2007] In Fig. 4.9 we show five  $\tau_{21}$ ,  $\omega_{1}$  plots for varying  $\omega_{2}$  with  $\tau_{22'}=0.^{7}$  The plots are the analogue to the most common multidimensional experiment of Transient Absorption spectroscopy, where the non-linear probe spectrum is plotted as a function of the pump-probe delay. For each plot, the  $\omega_{2}$  frequency is denoted by a vertical gray line. Each Wigner plot is scaled to its own dynamic range to emphasize the dependence on  $\omega_{2}$ . The dramatic line shape changes between positive and negative delays can be seen. This representation also highlights the asymmetric broadening of the  $\omega_{1}$  line shape near pulse overlap when  $\omega_{2}$  becomes non-resonant. Again, these features can resemble spectral diffusion even though our system is homogeneous.

# 4.4.5 Inhomogeneous broadening

<sup>&</sup>lt;sup>7</sup>See Supplementary Fig. S8 for Wigner plots for all  $\Gamma_{10}\Delta_t$  values.




Figure 4.10: 2D delay response for  $\Gamma_{10}\Delta_t=1$  with sample inhomogeneity. All pulses are tuned to exact resonance. The colors depict the signal amplitude. The black contour lines show signal purity, P (see Equation 4.25), with purity values denoted on each contour. The thick yellow line denotes the peak amplitude position that is used for 3PEPS analysis. The small plot above each 2D delay plot examines a  $\tau_{22'}$  slice of the delay response ( $\tau_{21}=0$ ). The plot shows the total signal (black), as well as the component time-orderings VI (orange), V (purple), III (teal, dashed), and I (teal, solid).

With the homogeneous system characterized, we can now consider the effect of inhomogeneity. For inhomogeneous systems, time-orderings III and V are enhanced because their final coherence will rephase to form a photon echo, whereas time-orderings I and VI will not. In delay space, this rephasing appears as a shift of signal to time-ordered regions III and V that persists for all population times. Fig. 4.10 shows the calculated spectra for relative dephasing rate  $\Gamma_{10}\Delta_t=1$  with a frequency broadening function of width  $\Delta_{\rm inhom}=0.441\Gamma_{10}$ . The inhomogeneity makes it easier to temporally isolate the rephasing pathways and harder to isolate the non-rephasing pathways, as shown by the purity contours.

A common metric of rephasing in delay space is the 3PEPS measurement. [Weiner1985, Fleming1998, Boeij1998, Salvador2003] In 3PEPS, one measures the signal as the first coherence time,  $\tau$ , is scanned across both rephasing and non-rephasing pathways while keeping population time, T, constant. The position of the peak is measured; a peak shifted away from  $\tau=0$  reflects the rephasing ability of the system. An inhomogeneous system will emit a photon echo in the rephasing pathway, enhancing signal in the rephasing time-ordering and creating the peak shift. In our 2D delay space, the  $\tau$  trace can be defined if we assume  $E_2$  and  $E_{2'}$  create the population (time-orderings V and VI). The trace runs parallel to the III-V time-ordering boundary (diagonal) if  $\tau_{22'} < 0$  and runs parallel to the IV-VI time-ordering boundary (horizontal) if  $\tau_{22'} > 0$ , and both intersect at  $\tau_{22'} = 0$ ; the  $-\tau_{21}$  value at this intersection is  $T.^8$  In our 2D delay plots (Fig. 4.7, Fig. 4.10), the peak shift is seen as the diagonal displacement of the signal peak from the  $\tau_{21} = 0$  vertical line. Fig. 4.10 highlights the peak shift profile as a function of population time with the yellow trace; it is easily verified that our static inhomogneous system exhibits a non-zero peak shift value for all population times.

The unanticipated feature of the 3PEPS analysis is the dependence on T. Even though our inhomogeneity is static, the peak shift is maximal at T=0 and dissipates as T increases, mimicking spectral diffusion. This dynamic arises from signal overlap with time-ordering III, which uses  $E_2$  and  $E_1$  as the first two interactions ,and merely reflects  $E_1$  and  $E_2$  temporal overlap. At T=0, the  $\tau$  trace gives two ways to make a rephasing pathway (time-orderings III and V) and only one way to make a non-rephasing pathway (time-ordering VI). This pathway asymmetry shifts signal away from  $\tau=0$  into the rephasing

 $<sup>^{8}</sup>$  See Supplementary Fig. S9 for an illustration of how 3PEPS shifts are measured from a 2D delay plot.

<sup>&</sup>lt;sup>9</sup> Supplementary Fig. S10 shows the 3PEPS measurements of all 12 combinations of  $\Gamma_{10}\Delta_t$  and  $\Delta_{inhom}$ , for every population delay surveyed.

direction. At large T (large  $\tau_{21}$ ), time-ordering III is not viable and pathway asymmetry disappears. Peak shifts imply inhomogeneity only when time-orderings V and III are minimally contaminated by each other i.e. at population times that exceed pulse overlap. This fact is easily illustrated by the dynamics of homogeneous system (Fig. 4.7); even though the homogeneous system cannot rephase, there is a non-zero peak shift near T=0. The contamination of the 3PEPS measurement at pulse overlap is well-known and is described in some studies,[DeBoeij1996, Agarwal2002] but the dependence of pulse and system properties on the distortion has not been investigated previously. Peak-shifting due to pulse overlap is less important when  $\omega_1 \neq \omega_2$  because time-ordering III is decoupled by detuning.




Figure 4.11: Same as Fig. 4.8, but each system has inhomogeneity ( $\Delta_{inhom}=0.441\Gamma_{10}$ ). Relative dephasing rates are  $\Gamma_{10}\Delta_t=0.5$  (red), 1.0 (green), and 2.0 (blue). In all plots  $\tau_{22'}=0$ . To ease comparison between different dephasing rates, the colored line shapes of all three systems are overlaid. Each 2D plot shows a single representative contour (half-maximum) for each  $\Gamma_{10}\Delta_t$  value. The colored histograms below each 2D frequency plot show the relative weights of each time-ordering for each 2D frequency plot. In contrast to Fig. 4.8, inhomogeneity makes the relative contributions of time-orderings V and VI unequal.

In frequency space, spectral elongation along the diagonal is the signature of inhomogeneous broadening. Fig. 4.11 shows the line shape changes of a Gaussian inhomogeneous distribution. All systems are broadened by a distribution proportional to their dephasing bandwidth. As expected, the sequence again shows a gradual broadening along the  $\omega_1$  axis, with a strong spectral correlation at early delays ( $\tau_{21} > 0$ ) for the more driven signals. The anti-diagonal width at early delays (e.g. Fig. 4.11,  $\tau_{21} = 2.0\Delta_t$ ) again depends on the pulse bandwidth and the monochromator slit width. At delay values that isolate time-orderings V and VI, however, the line shapes retain diagonal character, showing the characteristic balance of homogeneous and inhomogeneous width.

#### 4.5 Discussion

#### 4.5.1 An intuitive picture of pulse effects

Our chosen values of the relative dephasing time,  $\Gamma_{10}\Delta_t$ , describe experiments where neither the impulsive nor driven limit unilaterally applies. We have illustrated that in this intermediate regime, the multidimensional spectra contain attributes of both limits, and that it is possible to judge when these attributes apply. In our three-pulse experiment the second and third pulses time-gate coherences and populations produced by the previous pulse(s), and the monochromator frequency-gates the final coherence. Time-gating isolates different properties of the coherences and populations. Consequently, spectra evolve against delay. For any delay coordinate, one can develop qualitative line shape expectations by considering the following three principles:

- 1. When time-gating during the pulse, the system pins to the driving frequency with a buildup efficiency determined by resonance.
- 2. When time-gating after the pulse, the FID dominates the system response.
- 3. The emitted signal field contains both FID and driven components; the  $\omega_{\text{out}} = \omega_1$  component is isolated by the tracking monochromator.

<sup>&</sup>lt;sup>10</sup>As in Fig. 4.8, Fig. 4.11 shows only the contours at the half-maximum amplitude. See Supplementary Fig. S7 for all contours.

Fig. 4.4 illustrates principles 1 and 2 and Fig. 4.5 illustrates principle 2 and 3. Fig. 4.6 provides a detailed example of the relationship between these principles and the multidimensional line shape changes for different delay times.

The principles presented above apply to a single pathway. For rapidly dephasing systems it is difficult to achieve complete pathway discrimination, as shown in Fig. 4.7. In such situations the interference between pathways must be considered to predict the line shape. The relative weight of each pathway to the interference can be approximated by the extent of pulse overlap. The pathway weights exchange when scanning across pulse overlap, which creates the dramatic line shape changes observed in Figs. 4.8 and 4.11.

#### 4.5.2 Conditional validity of the driven limit

We have shown that the driven limit misses details of the line shape if  $\Gamma_{10}\Delta_t\approx 1$ , but we have also reasoned that in certain conditions the driven limit can approximate the response well (see principle 1). Here we examine the line shape at delay values that demonstrate this agreement. Fig. 4.12 compares the results of our numerical simulation (third column) with the driven limit expressions for populations where  $\Gamma_{11}\Delta_t=0$  (first column) or 1 (second column). The top and bottom rows compare the line shapes when  $(\tau_{22'},\tau_{21}=(0,0))$  and  $(0,-4\Delta_t)$ , respectively. The third column demonstrates the agreement between the driven limit approximations with the simulation by comparing the diagonal and anti-diagonal cross-sections of the 2D spectra.

Note the very sharp diagonal feature that appears for  $(\tau_{21}, \tau_{22'}) = (0,0)$  and  $\Gamma_{11} = 0$ ; this is due to population resonance in time-orderings I and III. This expression is inaccurate: the narrow resonance is only observed when pulse durations are much longer than the coherence time. A comparison of picosecond and femtosecond studies of quantum dot exciton line shapes (Yurs *et al.*[Yurs2011] and Kohler *et al.*[Kohler2014], respectively) demonstrates this difference well. The driven equation fails to reproduce our numerical simulations here because resonant excitation of the population is impulsive; the experiment time-gates only the rise time of the population, yet driven theory predicts the resonance to be vanishingly narrow ( $\Gamma_{11} = 0$ ). In light of this, one can approximate this time-gating effect by

substituting population lifetime with the pulse duration ( $\Gamma_{11}\Delta_t=1$ ), which gives good agreement with the numerical simulation (third column).

When  $\tau_{22'}=0$  and  $\tau_{21}<\Delta_t$ , signals can also be approximated by driven signal (Fig. 4.12 bottom row). Only time-orderings V and VI are relevant. The intermediate population resonance is still impulsive but it depends on  $\omega_{2'}-\omega_2$  which is not explored in our 2D frequency space.




Figure 4.12: Comparing approximate expressions of the 2D frequency response with the directly integrated response.  $\Gamma_{10}\Delta_t=1$ . The top row compares the 2D response of all time-orderings ( $\tau_{21}=0$ ) and the bottom row compares the response of time-orderings V and VI ( $\tau_{21}=-4\Delta_t$ ). First column: The driven limit response. Note the narrow diagonal resonance for  $\tau_{21}=0$ . Second column: Same as the first column, but with ad hoc substitution  $\Gamma_{11}=\Delta_t$ . Third column: The directly integrated response.

### 4.5.3 Extracting true material correlation




Figure 4.13: Temporal (3PEPS) and spectral (ellipticity) metrics of correlation and their relation to the true system inhomogeneity. The left column plots the relationship at pulse overlap (T=0) and the right column plots the relationship at a delay where driven correlations are removed ( $T=4\Delta_t$ ). For the ellipticity measurements,  $\tau_{22'}=0$ . In each case, the two metrics are plotted directly against system inhomogeneity (top and middle row) and against each other (bottom row). Colored lines guide the eyes for systems with equal relative dephasing rates ( $\Gamma_{10}\Delta_t$ , see upper legend), while the area of the data point marker indicates the relative inhomogeneity ( $\Delta_{\rm inhom}/\Gamma_{10}$ , see lower legend). Gray lines indicate contours of constant relative inhomogeneity (scatter points with the same area are connected).

We have shown that pulse effects mimic the qualitative signatures of inhomogeneity. Here we address how one can extract true system inhomogeneity in light of these effects. We focus on two ubiquitous metrics of inhomogeneity: 3PEPS for the time domain and ellipticity<sup>11</sup> for the frequency domain[Kwac2003, Okumura1999]. In the driven (impulsive) limit, ellipticity (3PEPS) corresponds to the frequency correlation function and uniquely extracts the inhomogeneity of the models presented here. In their respective limits, the metrics give values proportional to the inhomogeneity.

Fig. 4.13 shows the results of this characterization for all  $\Delta_{\rm inhom}$  and  $\Gamma_{10}\Delta_t$  values explored in this work. We study how the correlations between the two metrics depend on the relative dephasing rate,  $\Gamma_{10}\Delta_t$ , the absolute inhomogeneity,  $\Delta_{\rm inhom}/\Delta_\omega$ , the relative inhomogeneity  $\Delta_{\rm inhom}/\Gamma_{10}$ , and the population time delay. The top row shows the correlations of the  $\Delta_{\rm 3PEPS}/\Delta_t$  3PEPS metric that represents the normalized coherence delay time required to reach the peak intensity. The upper right graph shows the correlations for a population time delay of  $T=4\Delta_t$  that isolates the V and VI time-orderings. For this time delay, the  $\Delta_{\rm 3PEPS}/\Delta_t$  metric works well for all dephasing times of  $\Gamma_{10}\Delta_t$  when the relative inhomogeneity is  $\Delta_{\rm inhom}/\Delta_\omega \ll 1$ . It becomes independent of  $\Delta_{\rm inhom}/\Delta_\omega$  when  $\Delta_{\rm inhom}/\Delta_\omega > 1$ . This saturation results because the frequency bandwidth of the excitation pulses becomes smaller than the inhomogeneous width and only a portion of the inhomogeneous ensemble contributes to the 3PEPS experiment. [Weiner1985] The corresponding graph for T=0 shows a large peak shift occurs, even without inhomogeneity. In this case, the peak shift depends on pathway overlap, as discussed in Section 4.4.5.

The middle row in Fig. 4.13 shows the ellipticity dependence on the relative dephasing rate and inhomogeneity assuming the measurement is performed when the first two pulses are temporally overlapped  $(\tau_{22'}=0)$ . For a  $T=4\Delta_t$  population time, the ellipticity is proportional to the inhomogeneity until  $\Delta_{\rm inhom}/\Delta_\omega\ll 1$  where the excitation bandwidth is wide compared with the inhomogeneity. Unlike 3PEPS, saturation is not observed because pulse bandwidth does not limit the frequency range scanned. The 3PEPS and ellipticity metrics are therefore complementary since 3PEPS works well for  $\Delta_{\rm inhom}/\Delta_\omega\ll 1$  and ellipticity works well for  $\Delta_{\rm inhom}/\Delta_\omega\gg 1$ . When all pulses are temporally over-

There are many ways to characterize the ellipticity of a peak shape. We adopt the convention  $\mathcal{E} = \left(a^2 - b^2\right)/\left(a^2 + b^2\right)$ , where a is the diagonal width and b is the antidiagonal width.

The simulations for each value of the 3PEPS and ellipticity data in Fig. 4.13 appear in Supplementary Figs. S10-S12.

lapped at T=0, the ellipticity is only weakly dependent on the inhomogeneity and dephasing rate. The ellipticity is instead dominated by the dependence on the excitation pulse frequency differences of time-orderings I and III that become important at pulse overlap.

It is clear from the previous discussion that both metrics depend on the dephasing and inhomogeneity. The dephasing can be measured independently in the frequency or time domain, depending upon whether the dephasing is very fast or slow, respectively. In the mixed frequency/time domain, measurement of the dephasing becomes more difficult. One strategy to address this challenge is to use both the 3PEPS and ellipticity metrics. The bottom row in Fig. 4.13 plots 3PEPS against ellipticity to show how the relationship between the metrics changes for different amounts of dephasing and inhomogeneity. The anti-diagonal contours of constant relative inhomogeneity show that these metrics are complementary and can serve to extract the system correlation parameters.

Importantly, the metrics are uniquely mapped both in the presence and absence of pulse-induced effects (demonstrated by T=0 and  $T=4\Delta_t$ , respectively). The combined metrics can be used to determine correlation at T=0, but the correlation-inducing pulse effects give a mapping significantly different than at  $T=4\Delta_t$ . At T=0, 3PEPS is almost nonresponsive to inhomogeneity; instead, it is an almost independent characterization of the pure dephasing. In fact, the T=0 trace is equivalent to the original photon echo traces used to resolve pure dephasing rates.[Aartsma1976] Both metrics are offset due to the pulse overlap effects. Accordingly, the region to the left of homogeneous contour is non-physical, because it represents observed correlations that are less than that given by pulse overlap effects. If the metrics are measured as a function of T, the mapping gradually changes from the left figure to the right figure in accordance with the pulse overlap. Both metrics will show a decrease, even with static inhomogeneity. If a system has spectral diffusion, the mapping at late times will disagree with the mapping at early times; both ellipticity and 3PEPS will be smaller at later times than predicted by the change in mappings alone.

#### 4.6 Conclusion

This study provides a framework to describe and disentangle the influence of the excitation pulses in mixed-domain ultrafast spectroscopy. We analyzed the features of mixed domain spectroscopy through detailed simulations of MR-CMDS signals. When pulse durations are similar to coherence times, resolution is compromised by time-bandwidth uncertainty and the complex mixture of driven and FID response. The dimensionless quantity  $\Delta_t$  ( $\Gamma_f + i\kappa_f \Omega_{fx}$ ) captures the balance of driven and FID character in a single field-matter interaction. In the nonlinear experiment, with multiple field-matter interactions, this balance is also controlled by pulse delays and frequency-resolved detection. Our analysis shows how these effects can be intuitive.

The dynamic nature of pulse effects can lead to misleading changes to spectra when delays are changed. When delays separate pulses, the spectral line shapes of individual pathways qualitatively change because the delays isolate FID contributions and de-emphasize driven response. When delays are scanned across pulse overlap, the weights of individual pathways change, further changing the line shapes. In a real system, these changes would all be present in addition to actual dynamics and spectral changes of the material.

Finally, we find that, in either frequency or time domain, pulse effects mimic signatures of ultrafast inhomogeneity. Even homogeneous systems take on these signatures. For mixed domain experiments, pulse effects induce spectral ellipticity and photon echo signatures, even in homogeneous systems. Driven character gives rise to pathway overlap peak shifting in the 2D delay response, which artificially produces rephasing near pulse overlap. Driven character also produces resonances that depend on  $\omega_1 - \omega_2$  near pulse overlap. Determination of the homogeneous and inhomogeneous broadening at ultrashort times is only possible by performing correlation analysis in both the frequency and time domain.

# Part II

# **Instrumental Development**

### **Chapter 5**

### **Software**

Cutting-edge science increasingly relies on custom software. In their 2008 survey, Hannay et al. [7] demonstrated just how important software is to the modern scientist.

- ightarrow 84.3% of surveyed scientists state that developing scientific software is important or very important for their own research.
- ightarrow 91.2% of surveyed scientists state that using scientific software is important or very important for their own research.
- $\rightarrow$  On average, scientists spend approximately 40% of their work time using scientific software.
- $\rightarrow$  On average, scientists spend approximately 30% of their work time developing scientific software.

Despite the importance of software to science and scientists, most scientists are not familiar with basic software engineering concepts. This is in part due to the their general lack of formal training in programming and software development. Hannay et al. [7] found that over 90% of scientists learn software development through 'informal self study'. Indeed, I myself have never been formally trained in software development.

Software development in a scientific context poses unique challenges. Many traditional software development paradigms demand an upfront articulation of goals and requirements. This allows the developers to carefully design their software, even before a single line of code is written. In her seminal 2005

case study Segal [8] describes a collaboration between a team of researchers and a contracted team of software engineers. Ultimately

#### 5.1 Overview

In the Wright Group, PyCMDS replaces the old acquisition softwares 'ps control', written by Kent Meyer and 'Control for Lots of Research in Spectroscopy' written by Schuyler Kain.

### 5.2 Wright Tools

WrightTools is a software package at the heart of all work in the Wright Group.

### 5.3 PyCMDS

PyCMDS directly addresses the hardware during experiments.

#### 5.3.1 Overview

PyCMDS has, through software improvements alone, dramatically lessened scan times...

- $\rightarrow$  simultaneous motor motion
- $\rightarrow$  digital signal processing
- $\rightarrow$  ideal axis positions 5.3.2

#### 5.3.2 Ideal Axis Positions

Frequency domain multidimensional spectroscopy is a time-intensive process. A typical pixel takes between one-half second and three seconds to acquire. Depending on the exact hardware being scanned and signal being detected, this time may be mostly due to hardware motion or signal collection. Due to the curse of dimensionality, a typical three-dimensional CMDS experiment contains roughly 100,000 pixels. CMDS hardware is transiently-reliable, so speeding up experiments is a crucial component of unlocking ever larger dimensionalities and higher resolutions.

One obvious way to decrease the scan-time is to take fewer pixels. Traditionally, multidimensional scans are done with linearly arranged points in each axis—this is the simplest configuration to program into the acquisition software. Because signal features are often sparse or slowly varying (especially so in high-dimensional scans) linear stepping means that *most of the collected pixels* are duplicates or simply noise. A more intelligent choice of axis points can capture the same nonlinear spectrum in a fraction of the total pixel count.

An ideal distribution of pixels is linearized in *signal*, not coordinate. This means that every signal level (think of a contour in the N-dimensional case) has roughly the same number of pixels defining it. If some generic multidimensional signal goes between 0 and 1, one would want roughly 10% of the pixels to be between 0.9 and 1.0, 10% between 0.8 and 0.9 and so on. If the signal is sparse in the space explored (imagine a narrow two-dimensional Lorentzian in the center of a large 2D-Frequency scan) this would place the majority of the pixels near the narrow peak feature(s), with only a few of them defining the large (in axis space) low-signal floor. In contrast linear stepping would allocate the vast majority of the pixels in the low-signal 0.0 to 0.1 region, with only a few being used to capture the narrow peak feature. Of course, linearizing pixels in signal requires prior expectations about the shape of the multidimensional signal—linear stepping is still an appropriate choice for low-resolution "survey" scans.

CMDS scans often posses correlated features in the multidimensional space. In order to capture such features as cheaply as possible, one would want to define regions of increased pixel density along the correlated (diagonal) lineshape. As a concession to reasonable simplicity, our acquisition software

(PyCMDS) assumes that all scans constitute a regular array with-respect-to the scanned axes. We can acquire arbitrary points along each axis, but not for the multidimensional scan. This means that we cannot achieve strictly ideal pixel distributions for arbitrary datasets. Still, we can do much better than linear spacing.

Almost all CMDS lineshapes (in frequency and delay) can be described using just a few lineshape functions:

- $\rightarrow$  exponential
- $\rightarrow \ \mathsf{Gaussian}$
- $\rightarrow$  Lorentzian
- → bimolecular

Exponential and bimolecular dynamics fall out of simple first and second-order kinetics (I will ignore higher-order kinetics here). Gaussians come from our Gaussian pulse envelopes or from normally-distributed inhomogeneous broadening. The measured line-shapes are actually convolutions of the above. I will ignore the convolution except for a few illustrative special cases. More exotic lineshapes are possible in CMDS—quantum beating and breathing modes, for example—I will also ignore these. Derivations of the ideal pixel positions for each of these lineshapes appear below.

#### 5.3.3 Exponential

Simple exponential decays are typically used to describe population and coherence-level dynamics in CMDS. For some generic exponential signal S with time constant  $\tau$ ,

$$S(t) = e^{-\frac{t}{\tau}}. (5.1)$$

We can write the conjugate equation to 5.1, asking "what t do I need to get a certain signal level?":

$$\log(S) = -\frac{t}{\tau} \tag{5.2}$$

$$t = -\tau \log(S). \tag{5.3}$$

So to step linearly in t, my step size has to go as  $-\tau \log(S)$ .

We want to go linearly in signal, meaning that we want to divide S into even sections. If S goes from 0 to 1 and we choose to acquire N points,

$$t_n = -\tau \log \left(\frac{n}{N}\right). \tag{5.4}$$

Note that  $t_n$  starts at long times and approaches zero delay. So the first  $t_1$  is the smallest signal and  $t_N$  is the largest.

Now we can start to consider realistic cases, like where  $\tau$  is not quite known and where some other longer dynamics persist (manifested as a static offset). Since these values are not separable in a general system, I'll keep S normalized between 0 and 1.

$$S = (1-c)e^{-\frac{t}{\tau_{\text{actual}}}} + c \tag{5.5}$$

$$S_n = (1-c)e^{-\frac{-\tau_{\text{step}}\log\left(\frac{n}{N}\right)}{\tau_{\text{actual}}}} + c$$
 (5.6)

$$S_n = (1-c) e^{-\frac{\tau_{\text{step}}}{\tau_{\text{actual}}} \log \left(\frac{N}{n}\right)} + c$$
 (5.7)

$$S_n = (1-c)\left(\frac{N}{n}\right)^{-\frac{\tau_{\text{step}}}{\tau_{\text{actual}}}} + c \tag{5.8}$$

$$S_n = (1-c) \left(\frac{n}{N}\right)^{\frac{\tau_{\text{step}}}{\tau_{\text{actual}}}} + c \tag{5.9}$$

Gaussian

Lorentzian

Bimolecular

## 5.4 WrightSim

WrightSim does simulations.




Figure 5.1: TODO

# Chapter 6

# Instrumental Development

- 6.1 Hardware
- 6.1.1 Delay Stages
- 6.2 Signal Acquisition

Old boxcar: 300 ns window, 10 micosecond delay. Onset of saturation 2 V.

- 6.2.1 Digital Signal Processing
- 6.3 Artifacts and Noise
- 6.3.1 Scatter

Scatter is a complex microscopic process whereby light traveling through a material elastically changes its propagation direction. In CMDS we use propagation direction to isolate signal. Scattering samples

defeat this isolation step and allow some amount of excitation light to reach the detector. In homodynedetected 4WM experiments,

$$I_{\text{detected}} = |E_{4\text{WM}} + E_1 + E_2 + E_{2'}|^2 \tag{6.1}$$

Where *E* is the entire time-dependent complex electromagnetic field. When expanded, the intensity will be composed of diagonal and cross terms:

$$I_{\text{detected}} = \overline{(E_1 + E_2)} E_{2'} + (E_1 + E_2) \overline{E_{2'}} + |E_1 + E_2|^2 + (E_1 + E_2) \overline{E_{4WM}}$$

$$+ (E_1 + E_2) \overline{E_{4WM}} + \overline{E_{2'}} E_{4WM} + E_{2'} \overline{E_{4WM}} + |E_{4WM}|^2$$
(6.2)

A similar expression in the case of heterodyne-detected 4WM is derived by Brixner et al. [9]. The goal of any 'scatter rejection' processing procedure is to isolate  $|E_{4WM}|^2$  from the other terms.

#### **Abandon the Random Phase Approximation**

#### Interference Patterns in TrEE

TrEE is implicitly homodyne-detected. Scatter from excitation fields will interfere on the amplitude level with TrEE signal, causing interference patterns that beat in delay and frequency space. The pattern of beating will depend on which excitation field(s) reach(es) the detector, and the parameterization of delay space chosen.

First I focus on the interference patterns in 2D delay space where all excitation fields and the detection field are at the same frequency.

Here I derive the slopes of constant phase for the old delay space, where  $d1=\tau_{2'1}$  and  $d2=\tau_{21}$ . For simplicity, I take  $\tau_1$  to be 0, so that  $\tau_{21}\to\tau_2$  and  $\tau_{2'1}\to\tau_{2'}$ . The phase of signal is then

$$\Phi_{\text{sig}} = e^{-((\tau_{2'} - \tau_2)\omega)} \tag{6.3}$$




Figure 6.1: Numerically simulated interference patterns between scatter and TrEE for the old delay parametrization. Each column has scatter from a single excitation field. The top row shows the measured intensities, the bottom row shows the 2D Fourier transform, with the colorbar's dynamic range chosen to show the cross peaks.

The phase of each excitation field can also be written:

$$\Phi_1 = e^0 \tag{6.4}$$

$$\Phi_2 = e^{-\tau_2 \omega} \tag{6.5}$$

$$\Phi_{2'} = e^{-\tau_{2'}\omega} \tag{6.6}$$

The cross term between scatter and signal is the product of  $\Phi_{\text{sig}}$  and  $\Phi_{\text{scatter}}$ . The cross terms are:

$$\Delta_1 = \Phi_{\text{sig}} = e^{-((\tau_{2'} - \tau_2)\omega)}$$
 (6.7)

$$\Delta_2 = \Phi_{\text{sig}} e^{-\tau_2 \omega} = e^{-((\tau_{2'} - 2\tau_2)\omega)}$$
 (6.8)

$$\Delta_{2'} = \Phi_{\text{sig}} e^{-\tau_{2'}\omega} = e^{-\tau_2\omega} \tag{6.9}$$

Figure 6.1 presents numerical simulations of scatter interference as a visual aid. See Yurs 2011 [10].

Here I derive the slopes of constant phase for the current delay space, where d1 =  $\tau_{22'}$  and d2 =  $\tau_{21}$ . I take  $\tau_2$  to be 0, so that  $\tau_{22'} \to \tau_{2'}$  and  $\tau_{21} \to \tau_1$ . The phase of the signal is then

$$\Phi_{\text{sig}} = e^{-((\tau_{2'} + \tau_1)\omega)} \tag{6.10}$$

The phase of each excitation field can also be written:

$$\Phi_1 = e^{-\tau_1 \omega} \tag{6.11}$$

$$\Phi_2 = e^0 \tag{6.12}$$

$$\Phi_{2'} = e^{-\tau_{2'}\omega} \tag{6.13}$$

The cross term between scatter and signal is the product of  $\Phi_{\text{sig}}$  and  $\Phi_{\text{scatter}}$ . The cross terms are:

$$\Delta_1 = \Phi_{\mathsf{sig}} \mathsf{e}^{-\tau_1 \omega} \quad = \quad \mathsf{e}^{-\tau_{2'} \omega} \tag{6.14}$$

$$\Delta_2 = \Phi_{\text{sig}} = e^{-((\tau_2 + \tau_1)\omega)}$$
 (6.15)

$$\Delta_{2'} = \Phi_{\mathsf{sig}} \mathsf{e}^{-\tau_{2'}\omega} \quad = \quad \mathsf{e}^{-\tau_1\omega} \tag{6.16}$$




Figure 6.2: Numerically simulated interference patterns between scatter and TrEE for the current delay parametrization. Each column has scatter from a single excitation field. The top row shows the measured intensities, the bottom row shows the 2D Fourier transform, with the colorbar's dynamic range chosen to show the cross peaks.

Figure 6.2 presents numerical simulations of scatter interference for the current delay parameterization.

#### Instrumental Removal of Scatter

The effects of scatter can be entirely removed from CMDS signal by combining two relatively straight-forward instrumental techniques: *chopping* and *fibrillation*. Conceptually, chopping removes intensity-level offset terms and fibrillation removes amplitude-level interference terms. Both techniques work by modulating signal and scatter terms differently so that they may be separated after light collection.

Table 6.1: Four shot-types in a general phase shifted parallel modulation scheme. The 'other' category represents anything that doesn't depend on either chopper, including scatter from other excitation sources, background light, detector voltage offsets, etc.

We use the dual chopping scheme developed by Furuta, Fuyuki, and Wada [11] called 'phase shifted parallel modulation'. In this scheme, two excitation sources are chopped at 1/4 of the laser repetition rate (two pulses on, two pulses off). Very similar schemes are discussed by Augulis and Zigmantas [12] and Heisler et al. [13] for two-dimensional electronic spectroscopy. The two chop patterns are phase-shifted to make the four-pulse pattern represented in Table 6.1. In principle this chopping scheme can be achieved with a single judiciously placed mechanical chopper - this is one of the advantages of Furuta's scheme. Due to practical considerations we have generally used two choppers, one on each OPA. The key to phase shifted parallel modulation is that signal only appears when both of your chopped beams are passed. It is simple to show how signal can be separated through simple addition and subtraction

of the A, B, C, and D phases shown in Table 6.1. First, the components of each phase:

$$A = I_{\text{other}} \tag{6.17}$$

$$B = I_1 + I_{\text{other}} \tag{6.18}$$

$$C = I_{\text{signal}} + I_1 + I_2 + I_{\text{other}}$$
 (6.19)

$$D = I_2 + I_{\text{other}} \tag{6.20}$$

Grouping into difference pairs,

$$A - B = -I_1 \tag{6.21}$$

$$C - D = I_{\text{signal}} + I_1 \tag{6.22}$$

So:

$$A - B + C - D = I_{\text{signal}} \tag{6.23}$$

I have ignored amplitude-level interference terms in this treatment because they cannot be removed via any chopping strategy. Interference between signal and an excitation beam will only appear in 'C'-type shots, so it will not be removed in Equation 6.23. To remove such interference terms, you must *fibrillate* your excitation fields.

An alternative to dual chopping is single-chopping and 'leveling'... this technique was used prior to May 2016 in the Wright Group... 'leveling' and single-chopping is also used in some early 2DES work... [9].

Figure 6.3 shows the effects of dual chopping for some representative  $MoS_2$  TA data. Each subplot is a probe wigner, with the vertical grey line representing the pump energy. Note that the single chopper passes pump scatter, visible as a time-invariant increase in intensity when the probe and monochromator are near the pump energy. Dual chopping efficiently removes pump scatter, but at the cost of signal to noise for the same number of laser shots. Taking twice as many laser shots when dual chopping brings the signal to noise to at least as good as the original single chopping.

Fibrillation is the intentional randomization of excitation phase during an experiment. Because the




Figure 6.3: Comparison of single and dual chopping in a  $MoS_2$  transient absorption experiment. Note that this data has not been processed in any way - the colorbar represents changes in intensity seen by the detector. The grey line near 2 eV represents the pump energy. The inset labels are the number of laser shots taken and the chopping strategy used.

interference term depends on the phase of the excitation field relative to the signal, averaging over many shots with random phase will cause the interference term to approach zero. This is a well known strategy for removing unwanted interference terms [14, 15].

#### 6.3.2 Normalization of dual-chopped self-heterodyned signal

Shot-by-shot normalization is not trivial for these experiments. As in table above, with 1 as pump and 2 as probe.

Starting with  $\Delta I$  from 6.23, we can normalize by probe intensity to get the popular  $\Delta I/I$  representation. Using the names defined above:

$$\frac{\Delta I}{I} = \frac{A - B + C - D}{D - A} \tag{6.24}$$

Now consider the presence of excitation intensity monitors, indicated by subscripts PR for probe and PU for pump.

We can further normalize by the pump intensity by dividing the entire expression by  $C_{PU}$ :

$$\frac{\Delta I}{I} = \frac{A - B + C - D}{(D - A) * C_{PU}} \tag{6.25}$$

Now, substituting in BRAZARD formalism:

$$A = constant (6.26)$$

$$B = SI_{PU}^B(1+\delta_{PU}^B) \tag{6.27}$$

$$C = I_{PR}^{C}(1 + \delta_{PR}^{C}) + SI_{PU}^{C}(1 + \delta_{PR}^{C})$$
 (6.28)

$$D = I_{PR}^{D}(1 + \delta_{PR}^{D}) {(6.29)}$$

$$\frac{\Delta I}{I} = \frac{\langle A \rangle - \frac{\langle B_{PU} \rangle B}{B_{PU}} + \frac{\langle C_{PU} \rangle \langle C_{PR}C}{C_{PU}C_{PR}} - \frac{\langle D_{PR} \rangle D}{D_{PR}}}{\langle PR \rangle \langle PU \rangle}$$
(6.30)

- 6.4 Light Generation
- 6.4.1 Automated OPA Tuning
- 6.5 Optomechanics
- 6.5.1 Automated Neutral Density Wheels

# Part III

# **Applications**

Chapter 7

**PbSe** 

## Chapter 8

## MX2

We report the first coherent multidimensional spectroscopy study of a MoS<sub>2</sub> film. A four-layer sample of MoS<sub>2</sub> was synthesized on a silica substrate by a simplified sulfidation reaction and characterized by absorption and Raman spectroscopy, atomic force microscopy, and transmission electron microscopy. State-selective coherent multidimensional spectroscopy (CMDS) on the as-prepared MoS<sub>2</sub> film resolved the dynamics of a series of diagonal and cross-peak features involving the spin—orbit split A and B excitonic states and continuum states. The spectra are characterized by striped features that are similar to those observed in CMDS studies of quantum wells where the continuum states contribute strongly to the initial excitation of both the diagonal and cross-peak features, while the A and B excitonic states contributed strongly to the final output signal. The strong contribution from the continuum states to the initial excitation of both the diagonal and cross-peak features, while the A and B excitonic states contributed strongly to the final output signal. The strong contribution from the continuum states to the initial excitation shows that the continuum states are coupled to the A and B excitonic states and that fast intraband relaxation is occurring on a sub-70 fs time scale. A comparison of the CMDS excitation signal and the absorption spectrum shows that the relative importance of the continuum states is determined primarily by their absorption strength. Diagonal and cross-peak features decay with a 680 fs time constant characteristic of exciton recombination and/or trapping. The short time dynamics are complicated by coherent and partially coherent pathways that become important when the excitation pulses are temporally overlapped. In this region, the coherent dynamics create diagonal features involving both the excitonic states and continuum states, while the partially coherent pathways contribute to cross-peak features.

### 8.1 Introduction

Transition metal dichalcogenides (TMDCs), such as  $MoS_2$ , are layered semiconductors with strong spinorbit coupling, high charge mobility, and an indirect band gap that becomes direct for monolayers. [16, 17] The optical properties are dominated by the A and B excitonic transitions between two HOMO spin-orbit split valence bands and the lowest state of the conduction band at the K and K' valleys of the two-dimensional hexagonal Brillouin zone. [18] The spin and valley degrees of freedom are coupled in individual TMDC layers as a result of the strong spin-orbit coupling and the loss of inversion symmetry. The coupling suppresses spin and valley relaxation since both spin and valley must change in a transition. These unusual properties have motivated the development of TMDC monolayers for next-generation nano/optoelectronic devices as well as model systems for spintronics and valleytronics applications. [17, 19, 20]

Ultrafast dynamics of the MoS<sub>2</sub> A and B electronic states have been measured by pump-probe, transient absorption, and transient reflection spectroscopy. [21, 4, 22, 23, 24] The spectra contain A and B excitonic features that result from ground-state bleaching (GSB), stimulated emission (SE), and excited-state absorption (ESA) pathways. The excitons exhibit biexponential relaxation times of  $\approx$ 10–20 and  $\approx$ 350–650 fs, depending on the fluence and temperature. The dependence on excitation frequency has not been explored in previous ultrafast experiments on MoS<sub>2</sub>, but it has played a central role in understanding exciton cooling dynamics and exciton-phonon coupling in studies of quantum dots. [25]

Coherent multidimensional spectroscopy (CMDS) is a complementary four wave mixing (FWM) methodology that differs from pump-probe, transient absorption, and transient reflection methods. [20, 21, 4, 22, 24, 26, 23] Rather than measuring the intensity change of a probe beam caused by the state population changes induced by a pump beam, CMDS measures the intensity of a coherent output beam created by interactions with three excitation pulses. The interest in CMDS methods arise from their ability to remove inhomogeneous broadening, define interstate coupling, and resolve coherent and

incoherent dynamics. [27, 28, 29, 10, 30, 31, 32, 33, 34] CMDS typically requires interferometric phase stability between excitation pulses, so CMDS has been limited to materials with electronic states within the excitation-pulse bandwidth. Multiresonant CMDS is a particularly attractive method for the broader range of complex materials because it does not require interferometric stability and is able to use independently tunable excitation pulses over wide frequency ranges.

The multiresonant CMDS used in this work employs two independently tunable excitation beams with frequencies  $\omega_1$  and  $\omega_2$ . The  $\omega_2$  beam is split into two beams, denoted by  $\omega_2$  and  $\omega_2'$ . These three beams are focused onto the MoS<sub>2</sub> thin film at angles, creating an output beam in the phase-matched direction  $\mathbf{k}_{\text{out}} = \mathbf{k}_1 - \mathbf{k}_2 + \mathbf{k}_{2'}$  where  $\mathbf{k}$  is the wave vector for each beam and the subscripts label the excitation frequencies. Multidimensional spectra result from measuring the output intensity dependence on frequency and delay times.

Figure 8.1 introduces our conventions for representing multidimensional spectra. Figure 8.1b,d are simulated data. Figure 8.1a shows one of the six time orderings of the three excitation pulses where  $\tau_{22'}\equiv t_2-t_{2'}>0$  and  $\tau_{21}\equiv t_2-t_1<0$ ; that is, the  $\omega_{2'}$  pulse interacts first and the  $\omega_1$  pulse interacts last. Figure 8.1b illustrates the 2D delay-delay spectrum for all six time orderings when  $\omega_1$  and  $\omega_2$  are both resonant with the same state. The color denotes the output amplitude. Along the negative ordinate where  $\tau_{22'}=0$ , interactions with the  $\omega_2$  and  $\omega_{2'}$  pulses create a population that is probed by  $\omega_1$ . Similarly, along the negative absicissa where  $\tau_{21}=0$ , interactions with the  $\omega_2$  and  $\omega_1$  pulses create a population that is probed by  $\omega_{2'}$ . The decay along these axes measures the population relaxation dynamics. Note that these delay representations differ from previous publications by our group. [35] This paper specifically explores the dynamics along the ordinate where  $\tau_{22'}$  is zero and the  $\tau_{21}$  delay is changed.

Figure 8.1c depicts the A and B excitonic transitions between the spin-orbit split valence bands and the degenerate conduction band states of  $MoS_2$ . Figure 8.1d illustrates the 2D frequency-frequency spectrum when  $\omega_1$  and  $\omega_2$  are scanned over two narrow resonances. The spectrum contains diagonal and cross-peaks that we label according to the excitonic resonances AA, AB BA, and BB for illustrative purposes. The dynamics of the individual quantum states are best visualized by 2D frequency-delay plots, which combine the features seen in Figure 8.1b,d.




Figure 8.1: (a) Example delays of the  $\omega_1$ ,  $\omega_2$ , and  $\omega_{2'}$  excitation pulses. (b) Dependence of the output intensity on the  $\tau_{22'}$  and  $\tau_{21}$  time delays for  $\omega_1=\omega_2$ . The solid lines define the regions for the six different time orderings of the  $\omega_1$ ,  $\omega_2$ , and  $\omega_{2'}$  excitation pulses. We have developed a convention for numbering these time orderings, as shown. (c) Diagram of the band structure of MoS<sub>2</sub> at the K point. The A and B exciton transitions are shown. (d) Two dimensional frequency-frequency plot labeling two diagonal and cross-peak features for the A and B excitons.

This works reports the first multiresonant CMDS spectra of  $MoS_2$ . It includes the excitation frequency dependence of the A and B excitonic-state dynamics. These experiments provide a fundamental understanding of the multidimensional  $MoS_2$  spectra and a foundation for interpreting CMDS experiments on more complex TMDC heterostructures. The experimental spectra differ from the simple 2D spectrum shown in Figure 8.1d and those of earlier CMDS experiments with model systems. The line shape of the CMDS excitation spectrum closely matches the absorption spectrum, but the line shape of the output coherence is dominated by the A and B excitonic features. The difference arises from fast, < 70 fs intraband relaxation from the hot A and B excitons of the continuum to the band edge. A longer, 680 fs relaxation occurs because of trapping and/or exciton dynamics. [24] The intensity of the cross-peaks depends on the importance of state filling and intraband relaxation of hot A excitons as well as the presence of interband population trnasfer of the A and B exciton states.

### 8.2 Methods

MoS $_2$  thin films were prepared via a Mo film sulfidation reaction, similar to methods reported by Laskar et al. [36]. A 1 nm amount of Mo (Kurt J. Lesker, 99.95%) metal was electron-beam evaporated onto a fused silica substrate at a rate of 0.05 Å/s. The prepared Mo thin films were quickly transferred to the center of a 1-inch fused silica tub furnace equipped with gas flow controllers (see Figure 8.2) and purged with Ar. The temperature of the Mo substrate was increased to 900 °C over the course of 15 min, after which 200 mg of sulfur was evaporated into the reaction chamber. Sulfidation was carried out for 30 min, and the furnace was subsequently cooled to room temperature; then the reactor tube was returned to atmospheric pressure, and the MoS $_2$  thin film samples were collected. The MoS $_2$  samples were characterized and used for CMDS experiments with no further preparation.

 $MoS_2$  thin film absorption spectra were collected by a Shimadzu 2401PC ultraviolet-visible spectraphotometer. Raman and photoluminescence experiments were carried out in parallel using a Thermo DXR Raman microscope with a  $100 \times 0.9$  NA focusing objective and a 2.0 mW 532 nm excitation source. Raman/PL measurements were intentionally performed at an excitation power of <8.0 mW to prevent sample damage. [37] Contact-mode atomic foce microscopy was performed with an Agilent 5500 AFM.




Figure 8.2: Schemiatic of the synthetic setup used for Mo thin film sulfidation reactions.

 $MoS_2$  film thickness was determined by scratching the sample to provide a clean step-edge between the  $MoS_2$  film and the fused silica substrate. TEM samples were prepared following the method outlined by Shanmugam, Durcan, and Yu [38] using concentrated KOH in a 35 °C oil bath for 20 minutes. The delaminated  $MoS_2$  sample was removed from the basic solution, rinsed five times with DI water, and transferred to a Cu-mesh TEM grid. TEM experiments were performed on a FEI Titan aberration corrected (S)TEM under 200 kV accelerating voltage.

The coherent multidimensional spectroscopy system used a 35 fs seed pulse, centered at 800 nm and generated by a 1 kHz Tsunami Ti-sapphire oscillator. The seed was amplified by a Spitfire-Pro regenerative amplifier. The amplified output was split to pump two TOPAS-C collinear optical parametric amplifiers. OPA signal output was immediately frequency doubled with BBO crystals, providing two  $\approx$ 50 fs independently tunable pulses denoted  $\omega_1$  and  $\omega_2$  with frequencies ranging from 1.62 to 2.12 eV. Signal and idler were not filtered out, but played no role due to their low photon energy. Pulse  $\omega_2$  was split into pulses labeled  $\omega_2$  and  $\omega_{2'}$  to create a total of three excitation pulses.

In this experiment we use motorized OPAs which allow us to set the output color in software. OPA1 and OPA2 were used to create the  $\omega_1$  and  $\omega_2$  frequencies, respectively. In Figure 8.4 we compare the spectral envelope generated by the OPA at each set color. Negative detuning values correspond to regions of the envelope lower in energy than the corresponding set color. The colorbar allows for comparison between set color intensities. The fluence values reported correspond to the brightest set color for each OPA. A single trace of OPA2 output at set color = 1.95 eV can be found in Figure 8.7.

After passing through automated delay stages (Newport SMC100 actuators), all three beams were focused onto the sample surface by a 1 meter focal length spherical mirror in a distorted BOXCARS geometry to form a 630, 580, and 580  $\mu$ m FWHM spot sizes for  $\omega_1$ ,  $\omega_2$ , and  $\omega_{2'}$ , respectively.

Figure 8.5 represents delay corrections applied for each OPA. The corrections were experimentally determined using driven FWM output from fused silica. Corrections were approximately linear against photon energy, in agreement with the normal dispersion of transmissive optics inside our OPAs and between the OPAs and the sample. OPA2 required a relatively small correction along  $\tau_{22'}$  (middle subplot) to account for any dispersion experienced differently between the two split beams. OPA1 was not split and




Figure 8.3: (a) Mask. (b) 2D delay spectra at the BB diagonal ( $\omega_1=\omega_2\approx 1.95$  eV) for transmissive and reflective geometries. Transmissive signal is a mixture of MoS<sub>2</sub> signal and a large amount of driven signal from the substrate that only appears in the pulse overlap region. Reflective signal is representative of the pure MoS<sub>2</sub> response.




Figure 8.4: OPA outputs at each color explored.




Figure 8.5: Spectral delay correction.

therefore needed no such correction.

Figure 8.3a represents the to-scale mask that defines our distorted BOXCARS configuration. Relative to the center of the BOXCARS mask (small black dot),  $\omega_1$ ,  $\omega_2$ , and  $\omega_{2'}$  enter the sample at angles of 5.0, 1.5, and 1.0 degrees. Each is angled only along the vertical or horizontal dimension, as indicated in Figure 8.3a. This distortion allowed us to remove a large amount of unwanted  $\omega_2$  and  $\omega_{2'}$  photons from our signal path (Figure 8.3a red star).  $\omega_1$  photons were less efficiently rejected, as we show below. The center of the BOXCARS mask was brought into the sample at  $\approx$ 45 degrees. All three beams had S polarization. After reflection, the output beam was isolated using a series of apertures, spectrally resolved with a monochromator (spectral resolution 9 meV). and detected using a photomultiplier (RCA C31034A).

Our experimental setup allowed for the collection of both transmissive and reflective (epi-directional) FWM signal. The 2D delay spectra in Figure 8.3b show the presence of a large nonresonant contribution at the origin for the transmissive FWM signal and weaker signals from the  $MoS_2$  thin film at negative values of  $\tau_{21}$  and  $\tau_{22'}$ . The nonresonant contribution is much weaker than the signals from the film for the reflective signal id is the geometry chosen for this experiment. This discrimination between a film and the substrate was also seen in reflective and transmissive CARS microscopy experiments. [39]

Once measured, the FWM signal was sent through a four-stage workup process to create the data set shown here. This workup procedure is visualized in Figure 8.6. We use a chopper and boxcar in active background subtraction mode (averaging 100 laser shots) to extract the FWM signal from  $\omega_1$  and  $\omega_2$  scatter. We collect this differential signal (Figure 8.6b) in software with an additional 50 shots of averaging. In post-process we subtract  $\omega_2$  scatter and smooth the data using a 2D Kaiser window. Finally, we represent the homodyne collected data as (sig)<sup>1/2</sup> to make the dynamics and line widths comparable to heterodyne-collected techniques like absorbance and pump-probe spectra. Throughout this work, zero signal on the color bar is set to agree with the average rather than the minimum of noise. Values below zero due to measurement uncertainty underflow the color bar and are plotted in white. This is especially evident in lots such as +120 fs in Figure 8.14, where there is no real signal. IPython [40] and matplotlib [41] were important for data processing and plotting in this work.




Figure 8.6: Visualization of data collection and processing. With the exception of (c), each subsequent pane represents an additional processing step on top of previous processing. The color bar of each image is separate. (a) Voltages read by the detector at teach color combination. The large vertical feature is  $\omega_1$  scatter; the shape is indicative of the power curve of the OPA. MoS<sub>2</sub> response can be barely seen above this scatter. (b) Data after chopping and active background subtraction at the boxcar (100 shots). (c) The portion of chopped signal that is not material response. This portion is extracted by averaging several collections at very positive  $\tau_{21}$  values, where no material response is present due to the short coherence times of MoS<sub>2</sub> electronic states. The largest feature is  $\omega_2$  scatter. Cross-talk between digital-to-analog channels can also be seen as the negative portion that goes as  $\omega_1$  intensity. (d) Signal after (c) is subtracted. (e) Smoothed data. (f) Amplitude level (square root) data. This spectrum corresponds to that at 0 delay in Figure 8.9. Note that the color bar's range is different than in Figure 8.9.




Figure 8.7: Characterization of the few-layer  $MoS_2$  film studied in this work. Optical images of the  $MoS_2$  thin film on fused silica substrate in (a) transmission and (b) reflection. (c) Raman spectrum of the  $E^1_{2g}$  and  $A_{1g}$  vibrational modes. (d) High-resolution TEM image and its corresponding FFT shown in the inset. (e) Absorption (blue), photoluminescence (green), Gaussian fits to the A and B excitons, along with the residules betwen the fits and absorbance (dotted), A and B exciton centers (dotted) and representative excitation pulse shape (red).

#### 8.3 Results and discussion

The few-layer  $MoS_2$  thin film sample studied in this work was prepared on a transparent fused silica substrate by a simple sufidation reaction of a Mo thin film using a procedure modified from a recent report. [36] Figure 8.7a and b show the homogeneous deposition and surface smoothness of the sample over the centimeter-sized fused silica substrate, respectively. The Raman spectrum shows the  $E_{2g}^1$  and  $A_{1g}$  vibrational modes (Figure 8.7c) that are characteristic of  $MoS_2$ . jteLiSongLin2012a The transmission electron micrograph (TEM) in Figure 8.7d shows the lattice fringes of the film with an inset fast Fourier transform (FFT) of the TEM image indicative of the hexagonal crystal structure of the film corresponding to the 0001 plane of  $MoS_2$ . [42] The  $MoS_2$  film thickness was determined to be 2.66 nm by atomic force microscopy and corresponds to approximately four monolayers. Figure 8.73 shows the absorption and fluorescence spectrum of the film along with the A and B excitonic line shapes that were extracted from the absorption spectrum. A representative excitation pulse profile is also shown in red for comparison.

Extracting the exciton absorbance spectrum is complicated by the large "rising background" signal from other  $MoS_2$  bands. With this in mind, we fit the second derivative absorption spectrum to a sum of two second derivative Gaussians, as seen in Figure 8.8. Conceptually, this method can be thought of as maximizing the smoothness (as opposed to minimizing the amplitude) of the remainder between the fit and the absorption spectrum. The fit parameters can be found in the inset table in Figure 8.8. The Gaussians themselves and the remainder can be found in Figure 8.8.

The multiresonant CMDS experiment uses  $\approx$ 70 fs excitation pulses created by two independently tunable optical parametric amplifiers (OPAs). Automated delay stages and neutral density filters set the excitation time delays over all values of  $\tau_{21}$  with  $\tau_{22'}=0$  and the pulse fluence to 90  $\mu$ J/cm<sup>2</sup> (114  $\mu$ J/cm<sup>2</sup>) for the  $\omega_1$  ( $\omega_2$  and  $\omega_{2'}$ ) beam(s). Each pulse was focused onto the sample using a distorted BOXCARS configuration. [43] The FWM signal was spatially isolated and detected with a monochromator that tracks the output frequency so  $\omega_m=\omega_1$ . In order to compare the FWM spectra with the absorption spectrum, the signal has been defined as the square root of the measured FWM signal since FWM depends quadratically on the sample concentration and path length.




Figure 8.8: Extraction of excitonic features from absorbance spectrum. (a) Second derivative spectra of absorbance (black) and fit second derivative spectrum (green). Gaussian fit parameters are shown in the inset table. (b) Absorption curve (black), Gaussian fits (blue and red), and remainder (black dotted).




Figure 8.9: 2D frequency-frequency spectra of the MoS<sub>2</sub> sample in the epi configuration. In all spectra  $\tau_{22'}=0$  fs, while  $\tau_{21}$  is designated in the bottom-right corner of each spectral panel. The color bar defines the square root of the intensity normalized to the most intense feature in the series of spectra. The integration of the signal onto the  $\hbar\omega_1=\hbar\omega_m$  and  $\hbar\omega_2$  axes are represented ans the blue curves in the top and right side plots, respectively. The side plots also contain the absorbance spectrum (black line) to aid intepretation of the dynamics of the integrated 2D signals. The dashed lines mark the centers of the A and B excitons, as designated from the absorption spectrum.

The main set of data presented in this work is an  $\omega_1\omega_2\tau_{21}$  "movie" with  $\tau_{22\prime}=0$ . ?? shows representative 2D frequency-frequency slices from this movie at increasingly negative  $\tau_{21}$  times. Each 2D frequency spectrum contains side plots along both axes that compare the absorbance spectrum (black) to the projection of the integrated signal onto the axis (blue). Along  $\omega_1$  (which for negative  $\tau_{21}$  times acts as the "probe") we observe two peaks corresponding to the A and B excitons. In contrast, we see no well-defined excitonic peaks along the  $\omega_2$  "pump" axis. Instead, the signal amplitude increases toward bluer  $\omega_2$  values. The decrease in FWM above 2.05 eV is caused by a drop in the  $\omega_2$  OPA power.

Figures 8.10 and 8.11 show representative 2D frequency-delay slices from this movie, where the absicissa is the  $\omega_1$  or  $\omega_2$  frequency, respectively, the ordinate is the  $\tau_{21}$  delay time, and the solid bold lines represent five different  $\omega_2$  or  $\omega_1$  frequencies. The color bar is normalized to the brightest feature in each subplot. This normalization allows comparison of the time dependence of the line shapes, positions, and relative signal amplitudes along the  $\omega_1$  or  $\omega_2$  axis directly.

Each subplot in Figure 8.10 is similar to published pump-probe, transient absorption, multidimensional land transient reflection experiments that have measured the electronic dynamics of the A and B excitons. [20, 21, 4, 22, 23, 24, 26, 44] These previous experiments measure relaxation dynamics on the same  $\approx$ 400-600 fs time scale that is characteristic of Figure 8.10.

Our experiments also show how the spectral features change as a function of the  $\omega_2$  excitation frequency. The top to subplots of Figure 8.10 reflect the changes in the AA and BA features, while the bottom two subplots reflect the changes in the AB and BB features. The figure highlights the changes in the relative amplitude of the A and B features as a function of excitation frequency. Both the line shapes and the dynamics of the spectral features are very similar. Figure 8.11 is an excitation spectrum that shows that the dynamics of the spectral features do not depend strongly on the  $\omega_1$  frequency.

The spectral features in Figures 8.9, 8.10 and 8.11 depend on the quantum mechanical interference effects caused by the different pathways. Figure 8.12 shows all of the Liouville pathways required to understand the spectral features. [45, 46] These pathways correspond to the time orderings labeled V and VI in Figure 8.7b. The letters denote the density matrix elements,  $\rho_{ij}$ , where g representest the ground state and e, e' represent any excitonic state. Interaction with the temporally overlapped  $\omega_2$  and




Figure 8.10: Mixed  $\omega_1$ — $\tau_{21}$  time—frequency representations of the 3D data set at five ascending  $\omega_2$  excitation frequencies (solid black lines) showing the impact of the  $\omega_2$  excitation frequency on the  $\omega_1$  spectral line shape as a function of time. The A and B exciton energies are marked as dashed lines within each spectrum.




Figure 8.11: Mixed  $\omega_2$ — $\tau_{21}$  time—frequency representations of the 3D data set at five ascending  $\omega_1$  probe frequencies (solid black lines) showing the impact of the  $\omega_1$  excitation frequency on the  $\omega_2$  spectral line shape as a function of time. The A and B exciton energies are marked as dashed lines within each spectrum.




Figure 8.12: Liouville pathways for Figure 8.10. gg and ee designate ground- and excited-state populations, the eg, 2e,e, and e'+e,e represent the excitonic and biexcitonic output coherences, and the arrows are labeled with the frequencies or population transfer responsible for the transitions. e and e' represent either A or B excitonic states.

 $\omega_{2'}$  pulses excites the ee excited-state population and bleaches the ground-state population. Subsequent interaction with  $\omega_1$  creates the output coherences for the diagonal spectral features when  $\omega_1=\omega_2$  or the cross-peak features when  $\omega_1\neq\omega_2$ . The stimulated emission (SE) and ground-state bleaching (GSB) pathways create the eg or e'g output coherences from the ee and gg populations, respectively, while the excited-state absorption pathway creates the 2e,e or e'+e,e biexcitonic output coherences. Figure 8.12 also includes a population transfer pathway from the ee excited-state population to an e'e' population from which similar SE and ESA pathways occur. Since the ESA pathways destructively interfere with the SE and GSB pathways, the output singal depends on the differences between the pathways. Factors that change the biexcitonic output coherences such as the transition moments, state filling (Pauli blocking), frequency shifts, or dephasing rate changes will control the output signal. State filling and ground-state depletion are important factors for MoS<sub>2</sub> since the transitions excite specific electron and hole spin and valley states in individual layers.

The state-filling and ground-state bleaching effects on the diagonal and cross-peak features in Figure 8.9 depend on the spin and valley states in the output coherence. [19] The effects of these spins will disappear as the spin and valley states return to equilibrium. [47, 48, 49] If we assume spin relaxation is negligible, the FWM transitions that create the diagonal features involve either the A or B ESA transitions, so the resulting 2e,e state includes two spin-aligned conduction band electrons and valence band holes. Similarly, the cross-peak regions denoted by AB or BA in Figure 8.1d will have two transitions involving an A exciton, so the initial e'+e,e state will include antialigned spins. A quantitative treatment of the cancellation effects between the GSB, SE, and ESA pathways requries knowledge of the transition moments and state degeneracies and is beyond the scope of this paper. [50]

The most important characteristic of the experimental spectra is the contrast between the absence of well-resolved excitonic features that depend on  $\omega_2$  in Figures 8.9 and 8.11 and the well-defined excitonic features that depend on  $\omega_1$ . It is also important to note that the projections of the signal amplitude onto the  $\omega_2$  axis in Figure 8.9 match closely with the continuum features in the absorption spectrum and that the line shapes of the features along the  $\omega_2$  axis in Figure 8.11 do not change appreciably for different delay times or  $\omega_1$  values. It should be noted that the excitation pulse bandwidth (see Figure 8.7e) contributes to the absence of well-resolved A and B excitonic features along  $\omega_2$ . The similarity to

the continuum states in the absorption spectrum and the absence of a strong dependence on  $\omega_1$  show that the continuum states observed at higher  $\omega_2$  frequencies participate directly in creating the final output coherences and that their increasing importance reflects the increasing absorption strength of higher energy continuum states. In contrast, the features dependent on  $\omega_1$  in Figures 8.9 and 8.10 match the line shapes of the A and B excitonic resonances. Although the relative amplitudes of the spectral features depend on the  $\omega_2$  frequency, they do not depend strongly on the delay times. These characteristics show that hot A and B excitonic states undergo rapid intraband population relaxation over a <70 fs time scale set by excitation pulses to the A and B excitonic states excited by  $\omega_1$ .

A central feature of Figures 8.9 and 8.10 is that the AB region is much brighter than the BA region. This difference is suprising because simple models predict cross-peaks of equal amplitude, as depicted in Figure 8.1d. The symmetry in simple models arises because the AB and BA cross-peaks involve the same four transitions. The symmetry between AB and BA may be broken by material processes such as population relaxation and transfer, the output coherence dephasing rates, and the bleaching and state-filling effects of the valence and conduction band states involved in the transitions.

We believe that ultrafast intraband population transfer breaks the symmetry of AB and BA cross-peaks. For the BA peak, the interactions with  $\omega_2$  and  $\omega_{2'}$  generate only A excitons that do not relax on <70 fs time scales. For the AB peak,  $\omega_2$  and  $\omega_{2'}$  generate two kinds of excitons: (1) B excitons and (2) hot excitons in the A band. Relaxation to A may occur by interband transitions of B excitons or intraband transitions of hot A excitons. Either will lead to the GSB, SE, and ESA shown in the ee  $\rightarrow$  e'e' population transfer pathways of Figure 8.12. This relaxation must occur on the time scale of our pulse-width since the cross-peak asymmetry is observed even during temporal overlap. We believe that intraband relaxation of hot A excitons is the main factor in breaking the symmetry between AB and BA cross-peaks. Figure 8.10 shows that B  $\rightarrow$  A interband relaxation occurs on a longer time scale. The B/A ratio is higher when  $\omega_2$  is resonant with the B excitonic transition than when  $\omega_2$  is lower than the A exciton frequency (the top subplot). If population transfer of holes from the B to A valence bands occurred during temporal overlap, the B/A ratio would be independent of pump frequency at  $\tau_2 1 < 0$ .

Figure 8.13 shows the delay transients at the different frequencies shown in the 2D spectrum. The colors of the dots on the 2D frequency-frequency spectrum match the colors of the transients. The transients




Figure 8.13: Transients taken at the different  $\omega_1$  and  $\omega_2$  frequencies indicated by the colored markers on the 2D spectrum. The dynamics are assigned to a 680 fs fast time constant (black solid line) and a slow time constant represented as an unchanging offset over this timescale (black dashed line).

were taken with a smaller step size and a longer time scale than the delay space explored in the 3D data set. The transients are quite similar. Our data are consistent with both monomolecular biexponential and bimolecular kinetic models and cannot discriminate between them. We have fit the decay kinetics to a single exponential model with a time constant of 680 fs and an offset that represents the long time decay. The 680 fs decay is similar to previously published pump-probe and transient absorption experiments. [22, 23, 51]

The spectral features change quantitatively for delay times near temporal overlap. Figure 8.14 shows a series of 2D spectra for both positive and negative  $\tau_{21}$  delay times with  $\tau_{22'}=0$ . Each spectrum is normalized to its brightest feature. The spectra at positive  $\tau_{21}$  delay times become rapidly weaker as the delay times become more positive until the features vanish into the noise at +120 fs. The spectra also develop more diagonal character as the delay time moves from negative to positive values. The AB cross-peak is also a strong feature in the spectrum at early times.

The pulse overlap region is complicated by the multiple Liouville pathways that must be considered. Additionally, interference between scattered light from the  $\omega_1$  excitation beam and the output signal becomes a larger factor as the FWM signal decreases. Figure 8.15 shows the  $\omega_1$ ,  $\omega_2$ , and  $\omega_{2'}$  time ordered pathway that becomes and important consideration for positive  $au_{21}$  delay times. Since  $au_{22'}=0$ , the initial  $\omega_1$  pulse creates an excited-state coherence, while the subsequent  $\omega_2$  and  $\omega_{2'}$  pulses create the output coherence. The output signal is only important at short  $\tau_{21} > 0$  values because the initially excited coherence dephases very rapidly. When  $\omega_1 \neq \omega_2$ , the first two interactions create an e'e zero quantum coherence that also dephases rapidly. However, when  $\omega_1=\omega_2$ , the first two interactions create an ee, gg population difference that relaxes on linger time scales. The resulting signal will therefore appear as the diagonal feature in Figure 8.14 (e.g., see the +40 fs 2D spectrum). In addition to the diagonal feature in Figure 8.14, there is also a vertical feature when  $\omega_1$  is resonant with the A excitonic transition as well as the AB cross-peak. These features are attributed to the pathways in Figure 8.12. Although these pathways are depressed when  $\tau_{21} > 0$ , there is sufficient temporal overlap between the  $\omega_2$ ,  $\omega_{2'}$ , and  $\omega_1$  pulses to make their contribution comparable to those in Figure 8.15. More positie values of  $\tau_{21}$  emphasize the Figure 8.15 pathways over the Figure 8.12 pathways, accounting for the increasing percentage of diagonal character at increasingly positive delays.




Figure 8.14: 2D frequency-frquency spectra near zero  $\tau_{21}$  delay times. The signal amplitude is normalized to the brightest features in each spectrum.




Figure 8.15: Liouville pathways for the  $\omega_1$ ,  $\omega_2$ , and  $\omega_{2'}$  time ordering of pulse interactions. e and e' represent either A or B excitonic states.

### 8.4 Conclusions

This paper presents the first coherent multidimensional spectroscopy of  $MoS_2$  thin films. CMDS methods are related to the earlier ultrafast pump-probe and transient absorption methods since they all share bleaching, stimulated emission, Pauli blocking, and excited-state absorption pathways, but they differ in how these pathways define the spectra. In addition, CMDS methods have many additional pathways that become important when the coherence dephasing times are longer than the excitation pulse widths. In this work, the dephasing times are short so the pathways are identical to transient absorption. This work reports the first frequency-frequency-delay spectra of MX<sub>2</sub> samples. These spectra are complementary to previous work because they allow a direct comparison between the initially excited excitonic states and the states creating the final output coherence. The spectra show that the same hot A and B exciton continuum states that are observed in the absorption spectrum also dominate the CMDS excitation spectra. They also show that rapid, ¡70 fs intraband relaxation occurs to create the band-edge A and B excitonic features observed in the CMSD spectrum. The relative intensity of the diagonal peak features depends on the relative absorption strength of the A and B excitons. The relative intensity of cross-peak features in the 2D spectra depends on the excitation frequency. Excitation at or above the B exciton feature creates strong cross-peaks associated with hot A and B excitons that undergo ultrafast intraband population transfer. Excitation below the B excitonic feature creates a weak cross-peak indicating Ainduced B-state bleaching but at a lower signal level corresponding to the lower optical density at this energy. Population relaxation occurs over  $\approx$ 680 fs, either by transfer to traps or by bimolecular charge recombination.

These experiments provide the understanding of  $MoS_2$  coherent multidimensional spectra that will form the foundation required to measure the dynamical processes occurring in more complex  $MoS_2$  and other TMDC heterostructures with quantum-state resolution. The frequency domain based multiresonant CMDS methods described in this paper will play a central role in these measurements. They use longer, independently tunable pulses that provide state-selective excitation over a wide spectral range without the requirement for interferometric stability.

# Chapter 9

# PEDOT:PSS

## 9.1 Introduction

Poly(3,4-ethylenedioxythiophene)-poly(styrenesulfonate) (PEDOT:PSS) is a transparent, electrically conductive (up to  $4380 \text{ S cm}^{-1}$  [52]) polymer. It has found widespread use as a flexible, cheap alternative to inorganic transparent electrodes such as indium tin oxide.

As a polymer, PEDOT:PSS implicitly contains a large amount of structural inhomogeneity. On top of this, PEDOT:PSS is a two component material, composed of PEDOT (low molecular weight, p-doped, highly conductive) and PSS (high molecular-weight, insulating, stabilizing). These two components segment into domains of conductive and non-conductive material, leading to even more structural inhomogeneity. Nonlinear spectroscopy may be able to shed light on the microscopic environment of electronic states within PEDOT:PSS.

## 9.2 Background

Complex microstructure:

- 1. PEDOT oligomers (6—18-mers)
- 2. these oligomers  $\pi$ -stack to form small nanocrystalites, 3 to 14 oligomers for pristine films to as many as 13—14 oligomers for more conductive solvent treated films
- 3. nanocrystallites then arrange into globular conductive particles in a pancakge-like shape
- 4. these particles themselves are then linked via PSS-rich domains and assembled into nanofibril geometry akin to a string of pearls
- 5. nanofibrils interweave to form thin films, with PSS capping layer at surface

Prior spectroscopy (absorption anisotropy, X-ray scattering, condutivity).

Broad in the infrared due to midgap states created during doping from charge-induced lattice relaxations. These electronic perturbations arise from injected holes producing a quinoidal distortion spread over 4-5 monomers of the CP aromatic backbone, collectively called a polaron. Energetically favorable to be spin-silent bipolaron.

#### 9.3 Methods

PEDOT:PSS (Orgacon Dry, Sigma Aldrich) was dropcast onto a glass microscope slide at 1 mg/mL at a tilt to ensure homogeneous film formation. The sample was heated at 100  $^{\circ}$ C for  $\sim$ 15 min to evaporate water.

An ultrafast oscillator (Spectra-Physics Tsunami) was used to prepare  $\sim$ 35 fs seed pulses. These were amplified (Spectra-Physics Spitfire Pro XP, 1 kHz), split, and converted into 1300 nm 40 fs pulses using two separate optical parametric amplifiers (Light Conversion TOPAS-C): "OPA1" and "OPA2". Pulses from OPA2 were split again, for a total of three excitation pulses:  $\omega_1$ ,  $\omega_2$  and  $\omega_{2'}$ . These were passed through motorized (Newport MFA-CC) retroreflectors to control their relative arrival time ("delay") at the sample:  $\tau_{21} = \tau_2 - \tau_1$  and  $\tau_{22'} = \tau_2 - \tau_{2'}$ . The three excitation pulses were focused into the sample in a 1° right-angle isoceles triange, as in the BOXCARS configuration. [43] Each excitation beam was 67 nJ focused into a 375  $\mu$ m symmetric Gaussian mode for an intensity of 67  $\mu$ J/cm². A new beam, emitted coherently from the sample, was isolated with apertures and passed into a monochromator (HORIBA

Jobin Yvon MicroHR, 140 mm focal length) with a visible grating (500 nm blaze 300 groves per mm). The monochromator was set to pass all colors (0 nm, 250  $\mu$ m slits) to keep the measurement impulsive. Signal was detected using an InSb photodiode (Teledyne Judson J10D-M204-R01M-3C-SP28). Four wave mixing was isolated from excitation scatter using dual chopping and digital signal processing.

## 9.4 Transmittance and reflectance

Figure 9.1 shows the transmission, reflectance, and extinction spectrum of the thin film used in this work.




Figure 9.1: Thin film spectra. Transmission, reflectance, and extinction spectrum of the thin film used in this work. Extinction is  $\log_{10}$  (transmission).

## 9.5 Three-pulse echo spectroscopy

Two dimensional  $\tau_{21}$ ,  $\tau_{22'}$  scans were taken for two phase matching configurations: (1)  $k_{\text{out}} = k_1 - k_2 + k_{2'}$  (3PE) and (2)  $k_{\text{out}} = k_1 + k_2 - k_{2'}$  (3PE\*). The rephasing and nonrephasing pathways exchange their time dependance between these two configurations. Comparing both pathways, rephasing-induced peak shifts can be extracted as in 3PE. [CITE] All data was modeled using numerical integration of the Liouville-von Numann equation.

Continuously variable ND filters (THORLABS NDC-100C-4M, THORLABS NDL-10C-4) were used to ensure that all three excitation pulse powers were equal within measurement error.

Figure 9.2 diagrams the phase matching mask used in this set of experiments.




Figure 9.2: Phase matching mask used in this experiment. Each successive ring subtends 1 degree, such that the excitation pulses are each angled one degree relative to the mask center. The two stars mark the two output poyntings detected in this work.

Figure 9.3 shows the ten raw 2D delay-delay scans that comprise the primary dataset described in this section. The rows correspond to the two phase matching conditions, as labeled.




Figure 9.3: CAPTION TODO

#### 9.5.1 Assignment of zero delay

The absolute position of complete temporal overlap of the excitation pulses (zero delay) is a crucial step in determining the magnitude of the epeak shift and therefore the total rephasing ability of the material. The strategy for assigning zero delay relies upon the intrinsic symmetry of the two-dimensional delay space. Figure 9.4 labels the six time-orderings (TOs) of the three pulses that are possible with two delays. The TO labeling scheme follow from a convention first defined my Meyer, Wright and Thompson. [CITE] [CITE] first discussed how these TOs relate to traditional 3PE experiments. Briefly, spectral peak shifts into the rephasing TOs III and V when inhomogeneous broadening creates a photon echo in the III and V rephasing pathways colored orange in Figure 9.4. For both phase-matching conditions, there are two separate 3PE peak shift traces (represented as black arrows in Figure 9.4), yielding four different measurements of the photon echo. Since both 3PE and 3PE\* were measured using the same alignment on the same day, the zero delay position is identical for the four photon echo measurements. We focus on this signature when assigning zero delay—zero is correct only when all four peak shifts agree. Conceptually, this is the two-dimensional analogue to the traditional strategy of placing zero such that the two conjugate peak shifts (3PE and 3PE\*) agree. [CITE]

We found that the 3PEPS traces agree best when the data in Figure 9.3 is offset by 19 fs in  $\tau_{22'}$  and 4 fs in  $\tau_{21}$ . Figure 9.5 shows the 3PEPS traces after correcting for the zero delay value. The entire 3PEPS trace ( $\tau$  vs T) is show for regions I, III (purple and light green traces) and V, VI (yellow and light blue traces) for the [PHASE MATCHING EQUATIONS] phase matching conditions, respectively. Peak-shift magnitudes were found with Gaussian figs on the intensity level, in accordance with 3PEPS convention. [CITE] The bottom subplot of Figure 9.6 shows the agreement between the four traces for T > 50 fs where pulse-overlap effects become negligible. These pulse-overlap effects cause the 3PEPS at small T even without inhomogeneous broadening. [CITE] At long T, the average static 3PEPS is 2.5 fs.




Figure 9.4: CAPTION TODO




Figure 9.5: CAPTION TODO




Figure 9.6: CAPTION TODO




Figure 9.7: CAPTION TODO

There is a deviation of the TO I-III 3PEPS\* trace (green line) from the other traces. It is attributed to a combination of excitation pulse distortions and line shape differences between OPA1 and OPA2 (see Figure 9.1) and small errors in the zero delay correction. Figure 9.7 shows what the four 3PEPS traces would like like for different choices of zero-delay. The inset numbers in each subplot denote the offset (from chosen zero) in each delay axis.

#### **Numerical model**

We simulated the 3PEPS response of PEDOT:PSS through numerical integration of the Liouville-von Neumann Equation. Integration was performed on a homogeneous, three-level system with coherent dynamics described by

$$\frac{1}{T_2} = \frac{1}{2T_1} + \frac{1}{T_2^*},\tag{9.1}$$

where  $T_2$ ,  $T_1$  and  $T_2^*$  are the net dephasing, population relaxation, and pure dephasing rates, respectively. A three-level system was used because a two-level system cannot explain the population relaxation observed at long populations times, T. This slow delcay may be the same as the slowly decaying optical nonlinearities in PEDOT:PSS. [CITE] Inhomogeneity was incorporated by convolving the homogeneous repsonse with a Gaussian distribution function of width  $\Delta_{\rm inhom}$  and allowing the resultant polarization to interfere on the amplitude level. This strategy captures rephasing peak shifts and ensemble dephasing.

It is difficult to determine the coherence dephasing and the inhomogeneous broadening using 3PE if both factors are large. To extract  $T_2^*$  and  $\Delta_{\rm inhom}$ , we focused on two key components of the dataset, coherence duration and peak shift at large T. Since dephasing is very fast in PEDOT:PSS, we cannot directly respove an exponential free induction decay (FID). Instead, our model focuses on the FWHM of the  $\tau$  trace to determine the coherence duration. At T > 50 fs, the transient has a FWHM of  $\sim 80$  fs (intensity level). For comparison, our instrumental response is estimated to be 70-90 fs, depending on the exact value of our puse duration  $\Delta_t$  (35-45 fs FWHM, intensity level). An experimental peak shift of 2.5 fs was extracted using the strategy described above. Taken together, it is clear that both pure

dephasing and ensemble dephasing influence FWHM and peak shift so it is important to find valuee of  $T_2^*$  and  $\Delta_{\text{inhom}}$  that uniquely constrain the measured response.




Figure 9.8: CAPTION TODO

We simulated the  $\tau$  trance for a variety of  $\Delta_{\rm inhom}$  and  $T_2$  values. The results for  $\Delta_t=40$  fs are summarized in Figure 9.8. The lines of constant  $T_2$  span from  $\Delta_{\rm inhom}=0$  (green left ends of curves) to the limit  $\Delta_{\rm inhom}\to\infty$  (blue right ends of curves). The lines of constant  $T_2$  demonstrate that ensemble dephasing reduces the transient duration and introduces a peak shift. The influence of inhomogeneity on the observables vanishes as  $T_2\to\infty$ .

We preformed simulations analogus to those in Figure 9.8 for pulse durations longer and smaller than  $\Delta_t=40$  fs. Longer pulse durations create solutions that do not intersect our experimental point (see right-most subplot of Figure 9.8), but shorter pulse durations do. [TABLE] summarizes the coherence dephasing time and inomogeneous broadening values that best matches the experimental FWHM and inhomogeneous broadening value for  $\Delta_t=35,40$  and 45 fs. Clearly, there is no upper limit that can provide an upper limit for the inhomogeneous broadening.

Table 9.1: CAPTION TODO




Figure 9.9: CAPTION TODO

Our model system does ans excellent job of reproducing the entire 2D transient within measurement error (Figure 9.9). The most dramatic disagreement is in the upper right, where the experiment decays much slower than the simulation. Our system description does not account for signal contributions in TOs II and IV, where double quantum coherence resonances are important. In addition, excitation pulse shapes may cause such distortions. Regardless, these contributions do not affect our analysis.

Extremely fast (single fs) carrier scattering time constants have also been observed for PEDOT-base conductive films. [CITES]

### 9.6 Frequency-domain transient grating spectroscopy

This section describes preliminary, unpublished work accomplished on PEDOT:PSS.

Chapter 10

**Pyrite** 

# Chapter 11

# BiVO4

pass

Part IV

**Appendix** 

# Appendix A

# **Public**

### A.1 Chemical systems

Chemical systems are complex! They contain many molecules ( $10^{25}$  in a cup of coffee, 1 trillion in each human cell). These molecules have multiple interaction modes, both internal (intramolecular) and external (intermolecular). The reactivity of the system taken as a whole can be dominated by very rare but very important species, *e.g.* catalysts.

Despite this complexity, scientists have gotten very good at describing chemical systems through representations of dynamic equilibrium. In such situations, several key parameters emerge:

- concentration
- timescale (rate)
- lengthscale

#### A.1.1 Concentration

#### A.1.2 Timescale

#### A.1.3 Lengthscale

### A.2 Analytical chemistry

Traditionally, chemists have seen fit to divide themselves into four specializations: analytical, inorganic, organic, and physical. In recent years, materials chemistry and chemical biology have become specializations in their own right. This dissertation focuses on analytical chemistry.

Analytical chemists separate, identify, and quantify chemical systems. To do this, we build instruments that exploit physical properties of the chemical components:

- separation science (chromatography, electrophoresis)
- mass spectrometry
- electrochemistry
- microscopy
- spectroscopy

Spectroscopy is a family of strategies that exploit the interaction of chemical systems with light.

## A.3 Spectroscopy

Molecules respond to electric fields. Static electric fields cause charged molecules (ions) to move, as in electrophoresis and mass spectrometry. Oscillating electric fields, also known as light, can interact directly with the molecules themselves, driving transitions. However, these transitions can only be driven with the appropriate frequency of light (resonance). Different frequencies (colors) of light interact with different kinds of transitions, revealing different features of the molecule of interest.

#### A.3.1 Nonlinear spectroscopy

Spectroscopy is fantastic, but sometimes simple experiments don't reveal everything. Nonlinear spectroscopy uses multiple electric fields simultaniously, revealing even more information about the chemical system.

#### A.4 Instrumentation

To accomplish nonlinear spectroscopy, specialized light sources are needed:

- gigantic electric fields
- ultrafast time resoution
- tunable frequencies

#### A.4.1 LASER

These sources are made using Light Amplified by the Stimulated Emission of Radiation (LASER).

By keeping a wide range of colors in phase simulatniously, we are able to create truly ultrafast pulses of light. The work presented in this dissertation was primarily taken using a 35 fs 1 KHz system.

35 fs (35  $\times$  10<sup>15</sup> second) pulses are incredibly short:

$$\frac{\text{pulse duration (35 fs)}}{\text{time between pulses (1 ms)}} \approx \frac{5.75 \text{ months}}{\text{age of universe (13.7 billion years)}} \tag{A.1}$$

proportionally, our sample spends 6 months in the "sun" for every age of the unverse in the dark.

Because all of the energy within the pulse is compressed to such a short period of time, these pulses are

also incredibly powerful:

$$\frac{\text{energy per pulse (4 mJ)}}{\text{pulse duration (35 fs)}} \approx \frac{\text{US electricity generation}(5.43 \times 10^{11} \textit{W})}{5} \tag{A.2}$$

this laser outputs electric fields one fifth as powerful as total US electricity generation (2016).

#### A.4.2 OPA

# Appendix B

## **Procedures**

- **B.1** Aligining TOPAS-C
- **B.2** Aligning Spitfire PRO
- **B.3** Air Handling
- **B.4** Six Month Maintenance

## **B.5** Tuning MicroHR Monochromator

Visible Grating.

Align the HeNe as perpendicular as possible to the monochromator entrance slit.

Move the grating angle until the HeNe falls on the exit slit.

Shine a flashlight through the entrance slit and observe the colour on the exit slit: if white, then you

are at 0-order (0 nm), if red, then you are at 1st order (632.8 nm).

Go to 0-order, narrow the slits, and slowly adjust the angle until the HeNe is going through the exit slit.

Go to Jovin Yvon/utilities and find the motor configuration program.

In the Gratings tab, select the 1st grating (1200 line density) and hit Calibrate.

In theoretical wavelength, enter 0 nm.

In experimental wavelength, enter the wavelength you observe from the control program.

Hit set.

## Appendix C

## **Hardware**

In this chapter I collect some of the specific hardware contribution details that do not belong in the body of the dissertation.

### C.1 Adjustable periscopes

OPAs output horizontal or vertical polarizations according to which tuning process is used. Our experiments are opinionated about polarization, so some strategy for aligning polarization is necessary. In addition, it is useful to bring all excitation beams to the same height. To this end, I designed and constructed two adjustable periscopes. Each periscope is designed to bring OPA output to table height standard (5 inches) while either keeping or switching polarization. Both polarization configurations take the same path length, so source polarization can be switched without large changes to zero delay. All of this is done with just two (switched polarization) or three (kept polarization) reflections. A picture of these periscopes is shown in C.1.

While these periscopes are easy to align, their unique design means that it is not necessarily obvious what the correct strategy is. The following strategy will always converge:

1. use two "magic" apertures along the output beamline




Figure C.1: CAPTION TODO

155

2. in flipped polarization (two mirror configuration):

• use the stage (green X, Y) to align near aperture

• use the upper mirror (yellow TA, TB) to align far aperture

• iterate above

3. in kept polarization (three mirror configuration):

use stage X (green X) and upper mirror height (yellow TC) to align near aperture

• use lower mirror (pink SA, SB) to align far aperture

• iterate above

The kept polarization alignment is derivative of the fixed polarization alignment. One must ensure that

the fixed polarization is correctly aligned at all times.

Mirror B (aqua) is magnetically mounted to switch between polarization conditions. Ensure that the

lower turning mirror (pink) does not bump into mirror B (aqua) in polarization swtiching configuration.

The lower turning mirror is on a rail (pink SC). This rail is a rough adjust for the same degree of freedom

as pink SA. Adjust the rail only to ensure that the beam is roughly centered on the free aperture of the

turning mirror.

The first reflection is often accomplished using a wedge, as OPA output may be strong enough to

damage downstream optics. This optic can and should be replaced if more of the OPA output is desired

on the table (keeping damage thresholds in mind).

C.1.1 Wedge polarization preference

TODO: wedges will be more efficent at reflecting horizontal / vertical at 45 degrees

C.2 Automated transmissive filters

**TODO** 

## C.3 Electronics

TODO

## Appendix D

## **Errata**

In this appendix I list the errors I am currently aware of in my publications and associated work.

### D.1 Czech 2015

The following is an errata for Czech et al. [53], published in November 2015.

- Reference 13 is identical to reference 9.
- In the last paragraph of the introduction the sentence "The experimental spectra differ from the simple 2D spectrum shown in Figure 1d and those of earlier CMDS experiments with model systems" appears. This sentence cites references 6 through 10. Instead, it should cite references 15 through 20.
- In the last paragraph begining on page 12148, the text "Automated delay stages and neutral density filters set the excitation time delays over all values of  $\tau_{21}$  with  $\tau_{22}=0$ " appears. For the second  $\tau$ , the subscript should read 22', not 22.
- Caption of Figure 5 reads, in part: "showing the impact of the  $\omega_1$  excitation frequency on the  $\omega_1$  spectral line shape". This should instead read "showing the impact of the  $\omega_1$  excitation frequency on the  $\omega_2$  spectral line shape". The subscript on the last  $\omega$  should be a 2 and not a 1.

 $\bullet$  Figure 6 e'+e,e' should read e'+e,e and vice versa.

# **Appendix E**

# Colophon

This chapter lays out the technical aspects of this dissertation as a software and data product, including instructions for obtaining the source and regeneration of figures and documents.

# **Bibliography**

- [1] Daniel Kahneman. *Thinking, Fast and Slow.* Farrar, Straus and Giroux, 2013. ISBN: 9780374533557.
- [2] Duckhwan Lee and Andreas C. Albrecht. "A Unified View of Raman, Resonance Raman, and Fluorescence Spectroscopy (and their Analogues in Two-Photon Absorption". In: *Advances in infrared and Raman Spectroscopy*. Ed. by R. J. H. Clark and R. E. Hester. 1st ed. London; New York, 1985. Chap. 4, pp. 179–213. ISBN: 9780471906742.
- [3] Jacques Pankove. Optical Processes in Semiconductors. Dover Publications, 1975. ISBN: 9780486602752.
- [4] Nardeep Kumar, Jiaqi He, Dawei He, Yongsheng Wang, and Hui Zhao. "Charge carrier dynamics in bulk MoS2 crystal studied by transient absorption microscopy". In: *Journal of Applied Physics* 113.13 (2013), p. 133702. DOI: 10.1063/1.4799110.
- [5] Ji-xin Cheng, Andreas Volkmer, Lewis D Book, and X Sunney Xie. "An Epi-Detected Coherent Anti-Stokes Raman Scattering (E-CARS) Microscope with High Spectral Resolution and High Sensitivity". In: *The Journal of Physical Chemistry B* 105.7 (Feb. 2001), pp. 1277–1280. DOI: 10.1021/jp003774a.
- [6] Austin P. Spencer, Hebin Li, Steven T. Cundiff, and David M. Jonas. "Pulse Propagation Effects in Optical 2D Fourier-Transform Spectroscopy: Theory". In: *The Journal of Physical Chemistry A* 119.17 (2015), pp. 3936–3960. DOI: 10.1021/acs.jpca.5b00001.
- [7] Jo Erskine Hannay, Carolyn MacLeod, Janice Singer, Hans Petter Langtangen, Dietmar Pfahl, and Greg Wilson. "How do scientists develop and use scientific software?" In: 2009 ICSE Workshop on Software Engineering for Computational Science and Engineering. Institute of Electrical and Electronics Engineers (IEEE), May 2009. DOI: 10.1109/secse.2009.5069155.
- [8] Judith Segal. "When Software Engineers Met Research Scientists: A Case Study". In: *Empirical Software Engineering* 10.4 (Oct. 2005), pp. 517–536. DOI: 10.1007/s10664-005-3865-y.
- [9] Tobias Brixner, Tomas Mancal, Igor V. Stiopkin, and Graham R. Fleming. "Phase-stabilized two-dimensional electronic spectroscopy". In: *The Journal of Chemical Physics* 121.9 (2004), p. 4221. DOI: 10.1063/1.1776112.
- [10] Lena A Yurs, Stephen B. Block, Andrei V Pakoulev, Rachel S. Selinsky, Song Jin, and John Wright. "Multiresonant Coherent Multidimensional Electronic Spectroscopy of Colloidal PbSe Quantum Dots". In: *The Journal of Physical Chemistry C* 115.46 (Nov. 2011), pp. 22833–22844. DOI: 10.1021/jp207273x.
- [11] Koichi Furuta, Masanori Fuyuki, and Akihide Wada. "Cross-Term Selective, Two-Pulse Correlation Measurements by Phase-Shifted Parallel Modulation for Analysis of a Multi-Photon Process". In: *Applied Spectroscopy* 66.12 (Dec. 2012), pp. 1475–1479. DOI: 10.1366/12-06657.

- [12] Ramūnas Augulis and Donatas Zigmantas. "Two-dimensional electronic spectroscopy with double modulation lock-in detection: enhancement of sensitivity and noise resistance." In: *Optics express* 19.14 (June 2011), pp. 13126–13133. DOI: 10.1364/0E.19.013126.
- [13] Ismael A. Heisler, Roberta Moca, Franco V A Camargo, and Stephen R. Meech. "Two-dimensional electronic spectroscopy based on conventional optics and fast dual chopper data acquisition". In: *Review of Scientific Instruments* 85.6 (June 2014), p. 063103. DOI: 10.1063/1.4879822.
- [14] Ivan C. Spector, Courtney M. Olson, Christopher J. Huber, and Aaron M. Massari. "Simple fully reflective method of scatter reduction in 2D-IR spectroscopy". In: Optics Letters 40.8 (Apr. 2015), pp. 1850–1852. DOI: 10.1364/OL.40.001850.
- [15] Brian L. McClain, Ilya J. Finkelstein, and M. D. Fayer. "Vibrational echo experiments on red blood cells: Comparison of the dynamics of cytoplasmic and aqueous hemoglobin". In: *Chemical Physics Letters* 392.4-6 (July 2004), pp. 324–329. DOI: 10.1016/j.cplett.2004.05.080.
- [16] Qing Hua Wang, Kourosh Kalantar-Zadeh, Andras Kis, Jonathan N. Coleman, and Michael S. Strano. "Electronics and optoelectronics of two-dimensional transition metal dichalcogenides". In: *Nature Nanotechnology* 7.11 (Nov. 2012), pp. 699–712. DOI: 10.1038/nnano.2012.193. URL: https://doi.org/10.1038/nnano.2012.193.
- [17] Kin Fai Mak, Changgu Lee, James Hone, Jie Shan, and Tony F. Heinz. "Atomically ThinMoS2: A New Direct-Gap Semiconductor". In: *Physical Review Letters* 105.13 (Sept. 2010). DOI: 10.1103/physrevlett.105.136805.
- [18] Alejandro Molina-Sánchez, Davide Sangalli, Kerstin Hummer, Andrea Marini, and Ludger Wirtz. "Effect of spin-orbit interaction on the optical spectra of single-layer, double-layer, and bulk MoS2". In: *Physical Review B* 88.4 (July 2013). DOI: 10.1103/physrevb.88.045412.
- [19] Xiaodong Xu, Wang Yao, Di Xiao, and Tony F. Heinz. "Spin and pseudospins in layered transition metal dichalcogenides". In: *Nature Physics* 10.5 (Apr. 2014), pp. 343–350. DOI: 10.1038/ nphys2942.
- [20] Di Xiao, Gui-Bin Liu, Wanxiang Feng, Xiaodong Xu, and Wang Yao. "Coupled Spin and Valley Physics in Monolayers of MoS2 and Other Group-VI Dichalcogenides". In: *Physical Review Letters* 108.19 (May 2012). DOI: 10.1103/physrevlett.108.196802.
- [21] H. Fang, C. Battaglia, C. Carraro, S. Nemsak, B. Ozdol, J. S. Kang, H. A. Bechtel, S. B. Desai, F. Kronast, A. A. Unal, G. Conti, C. Conlon, G. K. Palsson, M. C. Martin, A. M. Minor, C. S. Fadley, E. Yablonovitch, R. Maboudian, and A. Javey. "Strong interlayer coupling in van der Waals heterostructures built from single-layer chalcogenides". In: *Proceedings of the National Academy of Sciences* 111.17 (Apr. 2014), pp. 6198–6202. DOI: 10.1073/pnas.1405435111.
- [22] Zhaogang Nie, Run Long, Linfeng Sun, Chung-Che Huang, Jun Zhang, Qihua Xiong, Daniel W. Hewak, Zexiang Shen, Oleg V. Prezhdo, and Zhi-Heng Loh. "Ultrafast Carrier Thermalization and Cooling Dynamics in Few-Layer MoS2". In: *ACS Nano* 8.10 (Oct. 2014), pp. 10931–10940. DOI: 10.1021/nn504760x.
- [23] Dezheng Sun, Yi Rao, Georg A. Reider, Gugang Chen, Yumeng You, Louis Brézin, Avetik R. Harutyunyan, and Tony F. Heinz. "Observation of Rapid Exciton–Exciton Annihilation in Monolayer Molybdenum Disulfide". In: Nano Letters 14.10 (Oct. 2014), pp. 5625–5629. DOI: 10.1021/n15021975.
- [24] Sangwan Sim, Jusang Park, Jeong-Gyu Song, Chihun In, Yun-Shik Lee, Hyungjun Kim, and Hyun-yong Choi. "Exciton dynamics in atomically thin MoS2: Interexcitonic interaction and broadening kinetics". In: *Physical Review B* 88.7 (Aug. 2013). DOI: 10.1103/physrevb.88.075434.

- [25] Patanjali Kambhampati. "Unraveling the Structure and Dynamics of Excitons in Semiconductor Quantum Dots". In: *Accounts of Chemical Research* 44.1 (Jan. 2011), pp. 1–13. DOI: 10.1021/ar1000428.
- [26] Kin Fai Mak, Keliang He, Jie Shan, and Tony F. Heinz. "Control of valley polarization in monolayer MoS2 by optical helicity". In: *Nature Nanotechnology* 7.8 (June 2012), pp. 494–498. DOI: 10. 1038/nnano.2012.96.
- [27] Steven T. Cundiff. "Coherent spectroscopy of semiconductors". In: *Optics Express* 16.7 (Mar. 2008), p. 4639. DOI: 10.1364/oe.16.004639.
- [28] Daniel B. Turner, Katherine W. Stone, Kenan Gundogdu, and Keith A. Nelson. "Three-dimensional electronic spectroscopy of excitons in GaAs quantum wells". In: *The Journal of Chemical Physics* 131.14 (Oct. 2009), p. 144510. DOI: 10.1063/1.3245964.
- [29] Daniel D. Kohler, Stephen B. Block, Schuyler Kain, Andrei V. Pakoulev, and John C. Wright. "Ultrafast Dynamics within the 1S Exciton Band of Colloidal PbSe Quantum Dots Using Multiresonant Coherent Multidimensional Spectroscopy". In: *The Journal of Physical Chemistry C* 118.9 (Mar. 2014), pp. 5020–5031. DOI: 10.1021/jp412058u. URL: https://doi.org/10.1021/jp412058u.
- [30] Graham B. Griffin, Sandrine Ithurria, Dmitriy S. Dolzhnikov, Alexander Linkin, Dmitri V. Talapin, and Gregory S. Engel. "Two-dimensional electronic spectroscopy of CdSe nanoparticles at very low pulse power". In: *The Journal of Chemical Physics* 138.1 (Jan. 2013), p. 014705. DOI: 10. 1063/1.4772465. URL: https://doi.org/10.1063/1.4772465.
- [31] Elad Harel, Sara M. Rupich, Richard D. Schaller, Dmitri V. Talapin, and Gregory S. Engel. "Measurement of electronic splitting in PbS quantum dots by two-dimensional nonlinear spectroscopy". In: Physical Review B 86.7 (Aug. 2012). DOI: 10.1103/physrevb.86.075412. URL: https://doi.org/10.1103/physrevb.86.075412.
- [32] S. T. Cundiff, M. Koch, W. H. Knox, J. Shah, and W. Stolz. "Optical Coherence in Semiconductors: Strong Emission Mediated by Nondegenerate Interactions". In: *Physical Review Letters* 77.6 (Aug. 1996), pp. 1107–1110. DOI: 10.1103/physrevlett.77.1107. URL: https://doi.org/10.1103/physrevlett.77.1107.
- [33] D. Birkedal, V. G. Lyssenko, J. M. Hvam, and K. El Sayed. "Continuum contribution to excitonic four-wave mixing due to interaction-induced nonlinearities". In: *Physical Review B* 54.20 (Nov. 1996), R14250–R14253. DOI: 10.1103/physrevb.54.r14250. URL: https://doi.org/10.1103/physrevb.54.r14250.
- [34] M. U. Wehner, D. Steinbach, and M. Wegener. "Ultrafast coherent transients due to exciton-continuum scattering in bulk GaAs". In: *Physical Review B* 54.8 (Aug. 1996), R5211–R5214. DOI: 10.1103/physrevb.54.r5211. URL: https://doi.org/10.1103/physrevb.54.r5211.
- [35] Andrei V. Pakoulev, Mark A. Rickard, Kent A. Meyer, Kathryn Kornau, Nathan A. Mathew, David E. Thompson, and John C. Wright. "Mixed Frequency/Time Domain Optical Analogues of Heteronuclear Multidimensional NMR". In: *The Journal of Physical Chemistry A* 110.10 (Mar. 2006), pp. 3352–3355. DOI: 10.1021/jp057339y. URL: https://doi.org/10.1021/jp057339y.
- [36] Masihhur R. Laskar, Lu Ma, Santhakumar Kannappan, Pil Sung Park, Sriram Krishnamoorthy, Digbijoy N. Nath, Wu Lu, Yiying Wu, and Siddharth Rajan. "Large area single crystal (0001) oriented MoS2". In: *Applied Physics Letters* 102.25 (June 2013), p. 252108. DOI: 10.1063/1.4811410. URL: https://doi.org/10.1063/1.4811410.

- [37] A. Castellanos-Gomez, M. Barkelid, A. M. Goossens, V. E. Calado, H. S. J. van der Zant, and G. A. Steele. "Laser-Thinning of MoS2: On Demand Generation of a Single-Layer Semiconductor". In: *Nano Letters* 12.6 (June 2012), pp. 3187–3192. DOI: 10.1021/nl301164v. URL: https://doi.org/10.1021/nl301164v.
- [38] Mariyappan Shanmugam, Chris A. Durcan, and Bin Yu. "Layered semiconductor molybdenum disulfide nanomembrane based Schottky-barrier solar cells". In: *Nanoscale* 4.23 (2012), p. 7399. DOI: 10.1039/c2nr32394j. URL: https://doi.org/10.1039/c2nr32394j.
- [39] Andreas Volkmer, Ji-Xin Cheng, and X. Sunney Xie. "Vibrational Imaging with High Sensitivity via Epidetected Coherent Anti-Stokes Raman Scattering Microscopy". In: *Physical Review Letters* 87.2 (June 2001). DOI: 10.1103/physrevlett.87.023901. URL: https://doi.org/10.1103/physrevlett.87.023901.
- [40] Fernando Perez and Brian E. Granger. "IPython: A System for Interactive Scientific Computing". In: Computing in Science & Engineering 9.3 (2007), pp. 21–29. DOI: 10.1109/mcse.2007.53. URL: https://doi.org/10.1109/mcse.2007.53.
- [41] John D. Hunter. "Matplotlib: A 2D Graphics Environment". In: Computing in Science & Engineering 9.3 (2007), pp. 90–95. DOI: 10.1109/mcse.2007.55. URL: https://doi.org/10.1109/mcse.2007.55.
- [42] Mark A. Lukowski, Andrew S. Daniel, Fei Meng, Audrey Forticaux, Linsen Li, and Song Jin. "Enhanced Hydrogen Evolution Catalysis from Chemically Exfoliated Metallic MoS2 Nanosheets". In: Journal of the American Chemical Society 135.28 (July 2013), pp. 10274–10277. DOI: 10.1021/ja404523s. URL: https://doi.org/10.1021/ja404523s.
- [43] Alan C. Eckbreth. "BOXCARS: Crossed-beam phase-matched CARS generation in gases". In: *Applied Physics Letters* 32.7 (Apr. 1978), pp. 421–423. DOI: 10.1063/1.90070.
- [44] Markus Thomalla and Helmut Tributsch. "Photosensitization of Nanostructured TiO2 with WS2 Quantum Sheets". In: *The Journal of Physical Chemistry B* 110.24 (June 2006), pp. 12167–12171. DOI: 10.1021/jp061371q. URL: https://doi.org/10.1021/jp061371q.
- [45] Andrei V. Pakoulev, Stephen B. Block, Lena A. Yurs, Nathan A. Mathew, Kathryn M. Kornau, and John C. Wright. "Multiply Resonant Coherent Multidimensional Spectroscopy: Implications for Materials Science". In: *The Journal of Physical Chemistry Letters* 1.5 (Mar. 2010), pp. 822–828. DOI: 10.1021/jz9003476. URL: https://doi.org/10.1021/jz9003476.
- [46] Andrei V. Pakoulev, Mark A. Rickard, Kathryn M. Kornau, Nathan A. Mathew, Lena A. Yurs, Stephen B. Block, and John C. Wright. "Mixed Frequency-/Time-Domain Coherent Multidimensional Spectroscopy: Research Tool or Potential Analytical Method?" In: Accounts of Chemical Research 42.9 (Sept. 2009), pp. 1310–1321. DOI: 10.1021/ar900032g. URL: https://doi.org/10.1021/ar900032g.
- [47] Hualing Zeng, Junfeng Dai, Wang Yao, Di Xiao, and Xiaodong Cui. "Valley polarization in MoS2 monolayers by optical pumping". In: *Nature Nanotechnology* 7.8 (June 2012), pp. 490–493. DOI: 10.1038/nnano.2012.95. URL: https://doi.org/10.1038/nnano.2012.95.
- [48] Bairen Zhu, Hualing Zeng, Junfeng Dai, and Xiaodong Cui. "The Study of Spin-Valley Coupling in Atomically Thin Group VI Transition Metal Dichalcogenides". In: Advanced Materials 26.31 (Apr. 2014), pp. 5504–5507. DOI: 10.1002/adma.201305367. URL: https://doi.org/10. 1002/adma.201305367.

- [49] Cong Mai, Andrew Barrette, Yifei Yu, Yuriy G. Semenov, Ki Wook Kim, Linyou Cao, and Kenan Gundogdu. "Many-Body Effects in Valleytronics: Direct Measurement of Valley Lifetimes in Single-Layer MoS2". In: *Nano Letters* 14.1 (Dec. 2013), pp. 202–206. DOI: 10.1021/n1403742j. URL: https://doi.org/10.1021/n1403742j.
- [50] Cathy Y. Wong and Gregory D. Scholes. "Using two-dimensional photon echo spectroscopy to probe the fine structure of the ground state biexciton of CdSe nanocrystals". In: *Journal of Luminescence* 131.3 (Mar. 2011), pp. 366–374. DOI: 10.1016/j.jlumin.2010.09.015. URL: https://doi.org/10.1016/j.jlumin.2010.09.015.
- [51] Callum J. Docherty, Patrick Parkinson, Hannah J. Joyce, Ming-Hui Chiu, Chang-Hsiao Chen, Ming-Yang Lee, Lain-Jong Li, Laura M. Herz, and Michael B. Johnston. "Ultrafast Transient Terahertz Conductivity of Monolayer MoS2 and WSe2 Grown by Chemical Vapor Deposition". In: ACS Nano 8.11 (Nov. 2014), pp. 11147–11153. DOI: 10.1021/nn5034746. URL: https://doi.org/10.1021/nn5034746.
- [52] Nara Kim, Seyoung Kee, Seoung Ho Lee, Byoung Hoon Lee, Yung Ho Kahng, Yong-Ryun Jo, Bong-Joong Kim, and Kwanghee Lee. "Highly Conductive PEDOT:PSS Nanofibrils Induced by Solution-Processed Crystallization". In: Advanced Materials 26.14 (Dec. 2013), pp. 2268–2272. DOI: 10.1002/adma.201304611.
- [53] Kyle J. Czech, Blaise J. Thompson, Schuyler Kain, Qi Ding, Melinda J. Shearer, Robert J. Hamers, Song Jin, and John C. Wright. "Measurement of Ultrafast Excitonic Dynamics of Few-Layer MoS2Using State-Selective Coherent Multidimensional Spectroscopy". In: ACS Nano 9.12 (Dec. 2015), pp. 12146–12157. DOI: 10.1021/acsnano.5b05198.